

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL021021\
 Data File : PL064753.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 10 Feb 2021 11:07
 Operator : AR\AJ
 Sample : PSTDCCC050
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 PSTDCCC050

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Feb 11 06:05:21 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL020821.M
 Quant Title : GC Extractables
 QLast Update : Tue Feb 09 04:41:29 2021
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.425	4.100	152.6E6	119.8E6	54.236	55.454
28)	SA Decachlor...	8.146	9.256	124.9E6	96559336	55.542	50.837
Target Compounds							
2)	A alpha-BHC	3.859	4.503	255.2E6	173.8E6	55.021	57.377
3)	MA gamma-BHC...	4.139	4.795	232.7E6	161.0E6	53.829	56.690
4)	MA Heptachlor	4.436	5.305	215.1E6	154.1E6	52.471	52.009
5)	MB Aldrin	4.680	5.615	203.3E6	136.1E6	54.588	55.494
6)	B beta-BHC	4.390	4.973	93223701	75893606	54.543	58.617
7)	B delta-BHC	4.591	5.194	227.6E6	144.3E6	56.312	53.405
8)	B Heptachlo...	5.121	6.010	189.7E6	127.8E6	54.643	57.944
9)	A Endosulfan I	5.456	6.372	173.6E6	114.9E6	54.142	54.780
10)	B gamma-Chl...	5.346	6.246	188.9E6	126.4E6	53.884	54.704
11)	B alpha-Chl...	5.404	6.318	186.4E6	125.7E6	53.829	54.599
12)	B 4,4'-DDE	5.570	6.475	171.4E6	109.5E6	54.760	54.018
13)	MA Dieldrin	5.698	6.632	179.8E6	115.4E6	54.824	53.979
14)	MA Endrin	5.953	6.854	146.5E6	87232866	48.763	48.254
15)	B Endosulfa...	6.226	7.069	157.2E6	99979491	54.598	53.058
16)	A 4,4'-DDD	6.084	6.973	143.5E6	92243921	56.665	53.209
17)	MA 4,4'-DDT	6.322	7.273	113.4E6	75554816	50.285	49.555
18)	B Endrin al...	6.395	7.195	130.4E6	85839520	53.479	52.715
19)	B Endosulfa...	6.606	7.418	152.3E6	96633707	53.701	53.697
20)	A Methoxychlor	6.869	7.734	64703339	44295108	49.970	49.001
21)	B Endrin ke...	7.093	7.897	163.3E6	103.5E6	56.946	54.118
22)	Mirex	7.274	8.347	115.2E6	74815297	54.525	53.823

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL021021\
Data File : PL064753.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 10 Feb 2021 11:07
Operator : AR\AJ
Sample : PSTDCCC050
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampleId :
PSTDCCC050

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Feb 11 06:05:21 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL020821.M
Quant Title : GC Extractables
QLast Update : Tue Feb 09 04:41:29 2021
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

