

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL021021\
 Data File : PL064760.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 10 Feb 2021 15:51
 Operator : AR\AJ
 Sample : M1347-01MS
 Misc :
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 OR-03-020921MS

Manual Integrations
 APPROVED

Ankita
 2/11/2021 8:58:47 AM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Feb 11 05:55:34 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL020821.M
 Quant Title : GC Extractables
 QLast Update : Tue Feb 09 04:41:29 2021
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.425	4.100	64515109	51216578	22.931	23.709
28)	SA Decachlor...	8.146	9.257	48538618	42643664	21.584	22.451
Target Compounds							
2)	A alpha-BHC	3.858	4.503	226.9E6	160.5E6	48.929	52.980
3)	MA gamma-BHC...	4.138	4.795	212.7E6	151.2E6	49.192	53.220
4)	MA Heptachlor	4.435	5.305	196.6E6	145.9E6	47.943	49.257
5)	MB Aldrin	4.680	5.616	170.7E6	123.5E6	45.847	50.378
6)	B beta-BHC	4.389	4.974	80701170	63181743	47.216	48.122
7)	B delta-BHC	4.590	5.195	196.9E6	127.4E6	48.712	47.138
8)	B Heptachlo...	5.121	6.010	167.9E6	108.7E6	48.356	49.254
9)	A Endosulfan I	5.456	6.373	157.8E6	104.8E6	49.203	49.987
10)	B gamma-Chl...	5.346	6.246	162.3E6	120.5E6	46.277	52.122
11)	B alpha-Chl...	5.403	6.319	160.7E6	115.1E6	46.393	50.007
12)	B 4,4'-DDE	5.570	6.475	144.4E6	101.8E6	46.145	50.218
13)	MA Dieldrin	5.696	6.633	178.1E6	100.2E6	54.308m	46.880
14)	MA Endrin	5.952	6.855	144.5E6	88591911	48.098	49.006
15)	B Endosulfa...	6.225	7.070	138.3E6	92970790	48.030	49.339
16)	A 4,4'-DDD	6.083	6.974	115.2E6	91049230	45.493	52.520
17)	MA 4,4'-DDT	6.321	7.273	107.6E6	78649294	47.720	51.585
18)	B Endrin al...	6.393	7.195	110.5E6	75902632	45.317	46.612
19)	B Endosulfa...	6.606	7.419	124.5E6	89120819	43.901	49.522
20)	A Methoxychlor	6.868	7.735	58327661	46270496	45.046	51.187
21)	B Endrin ke...	7.092	7.898	123.7E6	89539330	43.143	46.838
22)	Mirex	7.272	8.348	98042040	69755823	46.401m	50.183

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL021021\
Data File : PL064760.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 10 Feb 2021 15:51
Operator : AR\AJ
Sample : M1347-01MS
Misc :
ALS Vial : 13 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampled :
OR-03-020921MS

Manual Integrations
APPROVED

Ankita
2/11/2021 8:58:47 AM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Feb 11 05:55:34 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL020821.M
Quant Title : GC Extractables
QLast Update : Tue Feb 09 04:41:29 2021
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

