

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL021022\
 Data File : PL072593.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 10 Feb 2022 13:33
 Operator : AR\AJ
 Sample : N1449-01
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 20220040-NACOTE-2-COMPOSITE

Manual Integrations
 APPROVED

Reviewed By :Abdul Mirza 02/11/2022
 Supervised By :Ankita Jodhani 02/11/2022

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Feb 11 00:47:20 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL020722.M
 Quant Title : GC Extractables
 QLast Update : Tue Feb 08 05:14:10 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.556	5.492	7236892	24906942	15.030	14.419
28) SA Decachlor...	10.318	11.384	6123213	11291827	8.088	7.121m

Target Compounds

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL021022\
Data File : PL072593.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 10 Feb 2022 13:33
Operator : AR\AJ
Sample : N1449-01
Misc :
ALS Vial : 9 Sample Multiplier: 1



Instrument :
ECD_L
ClientSampleId :
20220040-NACOTE-2-COMPOSITE


Manual Integrations
APPROVED

Reviewed By : Abdul Mirza 02/11/2022
Supervised By : Ankita Jodhani 02/11/2022

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Feb 11 00:47:20 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL020722.M
Quant Title : GC Extractables
QLast Update : Tue Feb 08 05:14:10 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm






#1 Tetrachloro-m-xylene

R.T.: 4.556 min
Delta R.T.: -0.001 min
Response: 7236892
Conc: 15.03 ng/ml

Instrument :
ECD_L
ClientSampleId :
20220040-NACOTE-2-COMPOSITE


Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 02/11/2022
Supervised By :Ankita Jodhani 02/11/2022




#1 Tetrachloro-m-xylene

R.T.: 5.492 min
Delta R.T.: 0.000 min
Response: 24906942
Conc: 14.42 ng/ml



#28 Decachlorobiphenyl

R.T.: 10.318 min
Delta R.T.: 0.000 min
Response: 6123213
Conc: 8.09 ng/ml



#28 Decachlorobiphenyl

R.T.: 11.384 min
Delta R.T.: -0.002 min
Response: 11291827
Conc: 7.12 ng/ml m