

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL021221\
 Data File : PL064778.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 12 Feb 2021 10:45
 Operator : AR\AJ
 Sample : PSTDCCC050
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 PSTDCCC050

Manual Integrations
 APPROVED

Ankita
 2/15/2021 9:40:18 AM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Feb 12 13:35:22 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL020821.M
 Quant Title : GC Extractables
 QLast Update : Tue Feb 09 04:41:29 2021
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.433	4.100	141.5E6	114.2E6	50.305	52.871
28)	SA Decachlor...	8.159	9.262	103.8E6	84737732	46.161	44.613
Target Compounds							
2)	A alpha-BHC	3.867	4.503	237.9E6	168.1E6	51.300	55.504
3)	MA gamma-BHC...	4.148	4.795	215.5E6	148.7E6	49.853	52.365
4)	MA Heptachlor	4.446	5.307	203.3E6	146.4E6	49.578	49.412
5)	MB Aldrin	4.691	5.617	187.0E6	129.2E6	50.228	52.679
6)	B beta-BHC	4.400	4.975	87737409	64474294	51.333	49.189
7)	B delta-BHC	4.602	5.196	213.3E6	135.2E6	52.770	50.039
8)	B Heptachlo...	5.133	6.012	178.6E6	119.6E6	51.441	54.192
9)	A Endosulfan I	5.468	6.375	165.0E6	108.6E6	51.436	51.792
10)	B gamma-Chl...	5.358	6.247	181.5E6	120.3E6	51.749	52.072m
11)	B alpha-Chl...	5.417	6.320	177.7E6	119.2E6	51.307	51.780
12)	B 4,4'-DDE	5.583	6.477	160.5E6	103.9E6	51.281	51.210
13)	MA Dieldrin	5.711	6.635	169.6E6	108.2E6	51.706	50.622
14)	MA Endrin	5.965	6.857	137.2E6	91203806	45.686	50.450
15)	B Endosulfa...	6.238	7.073	149.7E6	93287836	51.983	49.507
16)	A 4,4'-DDD	6.096	6.977	137.4E6	86172271	54.235	49.707
17)	MA 4,4'-DDT	6.335	7.276	107.0E6	70502119	47.436	46.241
18)	B Endrin al...	6.407	7.198	122.9E6	80673310	50.409	49.542
19)	B Endosulfa...	6.619	7.422	144.2E6	90120168	50.842	50.078
20)	A Methoxychlor	6.882	7.738	58399771	40623299	45.102	44.939
21)	B Endrin ke...	7.106	7.901	152.9E6	96644259	53.326	50.554
22)	Mirex	7.287	8.351	107.6E6	68441492	50.944	49.238

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL021221\
Data File : PL064778.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 12 Feb 2021 10:45
Operator : AR\AJ
Sample : PSTDCCC050
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampleId :
PSTDCCC050

Manual Integrations
APPROVED

Ankita
2/15/2021 9:40:18 AM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Feb 12 13:35:22 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL020821.M
Quant Title : GC Extractables
QLast Update : Tue Feb 09 04:41:29 2021
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

