

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL021319\
 Data File : PL044751.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 13 Feb 2019 17:19
 Operator : AJ\SJ
 Sample : K1343-19
 Misc :
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 JC-02-021119-J

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Feb 14 00:17:12 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL020619.M
 Quant Title : GC Extractables
 QLast Update : Thu Feb 07 01:39:50 2019
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.365	3.985	57825669	19128209	8.126	8.142
28)	SA Decachlor...	8.052	9.047	66274254	21164478	9.087	9.556
Target Compounds							
12)	B 4,4'-DDE	5.485	6.323	4694268	1984331	0.541m	0.831 #



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL021319\
Data File : PL044751.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 13 Feb 2019 17:19
Operator : AJ\SJ
Sample : K1343-19
Misc :
ALS Vial : 19 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampleId :
JC-02-021119-J

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Feb 14 00:17:12 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL020619.M
Quant Title : GC Extractables
QLast Update : Thu Feb 07 01:39:50 2019
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x0.25µm






#1 Tetrachloro-m-xylene


R.T.: 3.365 min
Delta R.T.: 0.000 min
Response: 57825669
Conc: 8.13 ng/ml

Instrument :
ECD_L
ClientSampleId :
JC-02-021119-J




#1 Tetrachloro-m-xylene

R.T.: 3.985 min
Delta R.T.: 0.000 min
Response: 19128209
Conc: 8.14 ng/ml




#12 4,4'-DDE

R.T.: 5.485 min
Delta R.T.: -0.002 min
Response: 4694268
Conc: 0.54 ng/ml m



#12 4,4'-DDE


R.T.: 6.323 min
Delta R.T.: 0.002 min
Response: 1984331
Conc: 0.83 ng/ml



#28 Decachlorobiphenyl

R.T.: 8.052 min
Delta R.T.: 0.003 min
Response: 66274254
Conc: 9.09 ng/ml

Instrument :
ECD_L
ClientSampleId :
JC-02-021119-J



#28 Decachlorobiphenyl

R.T.: 9.047 min
Delta R.T.: 0.005 min
Response: 21164478
Conc: 9.56 ng/ml