

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL021319\
 Data File : PL044755.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 13 Feb 2019 18:15
 Operator : AJ\SJ
 Sample : K1343-27
 Misc :
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 JC-02-021119-N

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Feb 14 00:17:56 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL020619.M
 Quant Title : GC Extractables
 QLast Update : Thu Feb 07 01:39:50 2019
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.365	3.985	64822466	21467444	9.110	9.138
28)	SA Decachlor...	8.051	9.046	72677384	23089708	9.965	10.425
Target Compounds							
11)	B alpha-Chl...	5.319	6.165	2520604	1759296	0.264m	0.635m#



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL021319\
Data File : PL044755.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 13 Feb 2019 18:15
Operator : AJ\SJ
Sample : K1343-27
Misc :
ALS Vial : 23 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampleId :
JC-02-021119-N

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Feb 14 00:17:56 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL020619.M
Quant Title : GC Extractables
QLast Update : Thu Feb 07 01:39:50 2019
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x0.25µm






#1 Tetrachloro-m-xylene


R.T.: 3.365 min
Delta R.T.: 0.000 min
Response: 64822466
Conc: 9.11 ng/ml

Instrument :
ECD_L
ClientSampleId :
JC-02-021119-N




#1 Tetrachloro-m-xylene

R.T.: 3.985 min
Delta R.T.: 0.000 min
Response: 21467444
Conc: 9.14 ng/ml




#11 alpha-Chlordane

R.T.: 5.319 min
Delta R.T.: 0.000 min
Response: 2520604
Conc: 0.26 ng/ml m



#11 alpha-Chlordane


R.T.: 6.165 min
Delta R.T.: 0.002 min
Response: 1759296
Conc: 0.63 ng/ml m



#28 Decachlorobiphenyl

R.T.: 8.051 min
Delta R.T.: 0.002 min
Response: 72677384
Conc: 9.96 ng/ml

Instrument :
ECD_L
ClientSampleId :
JC-02-021119-N



#28 Decachlorobiphenyl

R.T.: 9.046 min
Delta R.T.: 0.003 min
Response: 23089708
Conc: 10.42 ng/ml