

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL021325\
 Data File : PL094176.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 13 Feb 2025 17:22
 Operator : AR\AJ
 Sample : PSTDCCC050
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 PSTDCCC050

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Feb 14 00:57:42 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL012125.M
 Quant Title : GC Extractables
 QLast Update : Tue Jan 21 14:58:05 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.538	2.774	127.9E6	163.4E6	47.482	50.068
28) SA Decachlor...	9.056	7.909	90843373	168.0E6	43.426	47.951
Target Compounds						
2) A alpha-BHC	3.994	3.276	186.0E6	247.3E6	48.515	50.586
3) MA gamma-BHC...	4.326	3.606	177.4E6	236.5E6	48.159	49.891
4) MA Heptachlor	4.915	3.944	157.5E6	224.2E6	48.069	48.163
5) MB Aldrin	5.256	4.224	160.0E6	227.4E6	48.902	49.860
6) B beta-BHC	4.525	3.907	76227242	98448751	47.425	49.288
7) B delta-BHC	4.772	4.135	169.7E6	236.6E6	48.407	49.787
8) B Heptachlo...	5.683	4.726	140.4E6	206.8E6	47.217	49.477
9) A Endosulfan I	6.069	5.096	124.7E6	187.6E6	47.193	48.396
10) B gamma-Chl...	5.939	4.976	132.7E6	211.4E6	47.595	49.884
11) B alpha-Chl...	6.018	5.040	132.1E6	207.0E6	47.382	49.446
12) B 4,4'-DDE	6.192	5.229	122.6E6	205.9E6	50.338	51.361
13) MA Dieldrin	6.344	5.360	128.7E6	211.2E6	46.351	49.164
14) MA Endrin	6.574	5.636	105.7E6	175.1E6	45.081	47.414
15) B Endosulfa...	6.794	5.931	108.1E6	183.6E6	44.860	49.583
16) A 4,4'-DDD	6.709	5.784	99430111	174.6E6	52.316	55.307
17) MA 4,4'-DDT	7.024	6.034	88209584	154.2E6	44.730	47.383
18) B Endrin al...	6.924	6.110	88052784	146.9E6	45.293	48.236
19) B Endosulfa...	7.159	6.333	100.6E6	175.4E6	44.443	49.172
20) A Methoxychlor	7.500	6.609	47633803	83453719	45.653	46.671
21) B Endrin ke...	7.644	6.838	115.6E6	213.1E6	45.839	50.801
22) Mirex	8.117	7.018	90147434	163.3E6	43.288	48.288

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL021325\
Data File : PL094176.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 13 Feb 2025 17:22
Operator : AR\AJ
Sample : PSTDCCC050
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampleId :
PSTDCCC050

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Feb 14 00:57:42 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL012125.M
Quant Title : GC Extractables
QLast Update : Tue Jan 21 14:58:05 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x0.25µm

