

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL021422\
 Data File : PL072689.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 15 Feb 2022 07:12
 Operator : AR\AJ
 Sample : PSTDCCC050
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 PSTDCCC050

Manual Integrations
 APPROVED

Reviewed By : Abdul Mirza 02/15/2022
 Supervised By : Ankita Jodhani 02/15/2022

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Feb 15 07:51:03 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL020722.M
 Quant Title : GC Extractables
 QLast Update : Tue Feb 08 05:14:10 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.555	5.490	23218585	79257852	48.221	45.883
28)	SA Decachlor...	10.312	11.380	33325411	77340920	44.020	48.772
Target Compounds							
2)	A alpha-BHC	5.261	6.071	32878603	116.0E6	49.865	48.444
3)	MA gamma-BHC...	5.684	6.467	32670011	108.5E6	51.982	47.105
4)	MA Heptachlor	6.089	7.113	35993709	105.2E6	49.311	46.200
5)	MB Aldrin	6.412	7.482	32885023	103.1E6	50.933	48.250
6)	B beta-BHC	6.068	6.728	11708375	48598831	47.251m	47.673
7)	B delta-BHC	6.326	6.993	33559463	109.2E6	50.685	47.646
8)	B Heptachlo...	6.990	7.949	32356667	92954564	51.323	48.181m
9)	A Endosulfan I	7.392	8.355	30230212	84902478	51.671	48.530
10)	B gamma-Chl...	7.267	8.219	32680160	93792046	51.590	48.490m
11)	B alpha-Chl...	7.336	8.300	32937323	93723731	51.725	49.329
12)	B 4,4'-DDE	7.551	8.491	31598300	87491561	51.996	46.810
13)	MA Dieldrin	7.676	8.645	32608872	85971707	51.637m	46.131
14)	MA Endrin	7.969	8.884	27907145	71261265	47.639	45.227
15)	B Endosulfa...	8.281	9.120	30221155	75352611	49.673	44.817
16)	A 4,4'-DDD	8.137	9.032	26631806	69185993	55.357	49.997
17)	MA 4,4'-DDT	8.397	9.348	23824019	62634417	44.276	40.915
18)	B Endrin al...	8.470	9.255	23049370	56507556	52.880	50.208
19)	B Endosulfa...	8.701	9.493	29683097	71147339	49.879	47.559
20)	A Methoxychlor	8.993	9.835	15277721	32558920	43.449	40.213
21)	B Endrin ke...	9.218	9.989	34736937	75635568	50.543	47.471
22)	Mirex	9.400	10.451	29145307	54651732	49.128	46.556

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL021422\
 Data File : PL072689.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 15 Feb 2022 07:12
 Operator : AR\AJ
 Sample : PSTDCCC050
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 PSTDCCC050

Manual Integrations APPROVED

Reviewed By :Abdul Mirza 02/15/2022
 Supervised By :Ankita Jodhani 02/15/2022

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Feb 15 07:51:03 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL020722.M
 Quant Title : GC Extractables
 QLast Update : Tue Feb 08 05:14:10 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

