

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL021521\
 Data File : PL064795.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 15 Feb 2021 09:39
 Operator : AR\AJ
 Sample : PSTDCCC050
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 PSTDCCC050

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Feb 15 14:41:07 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL020821.M
 Quant Title : GC Extractables
 QLast Update : Tue Feb 09 04:41:29 2021
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.425	4.099	141.5E6	115.8E6	50.289	53.618
28)	SA Decachlor...	8.145	9.256	105.1E6	87491282	46.715	46.063
Target Compounds							
2)	A alpha-BHC	3.858	4.502	243.6E6	172.3E6	52.538	56.889
3)	MA gamma-BHC...	4.138	4.794	221.3E6	158.7E6	51.194	55.866
4)	MA Heptachlor	4.435	5.305	201.0E6	151.4E6	49.034	51.092
5)	MB Aldrin	4.680	5.615	188.9E6	134.2E6	50.740	54.710
6)	B beta-BHC	4.389	4.973	86968020	71766089	50.883	55.209
7)	B delta-BHC	4.590	5.194	213.0E6	141.9E6	52.682	52.525
8)	B Heptachlo...	5.120	6.009	176.7E6	121.4E6	50.884	55.005
9)	A Endosulfan I	5.455	6.371	161.7E6	112.5E6	50.412	53.624
10)	B gamma-Chl...	5.346	6.245	177.8E6	124.2E6	50.699	53.742
11)	B alpha-Chl...	5.403	6.317	174.4E6	123.5E6	50.349	53.662
12)	B 4,4'-DDE	5.569	6.474	160.0E6	108.6E6	51.113	53.534
13)	MA Dieldrin	5.697	6.633	169.3E6	113.7E6	51.619	53.210
14)	MA Endrin	5.951	6.854	139.3E6	86769473	46.369	47.998
15)	B Endosulfa...	6.224	7.069	149.4E6	97912154	51.880	51.961
16)	A 4,4'-DDD	6.082	6.973	139.0E6	92836357	54.898	53.551
17)	MA 4,4'-DDT	6.320	7.273	106.1E6	72777597	47.064	47.734
18)	B Endrin al...	6.392	7.195	124.4E6	84409952	51.022	51.837
19)	B Endosulfa...	6.605	7.418	146.1E6	95152447	51.512	52.874
20)	A Methoxychlor	6.868	7.734	57173457	42036055	44.155	46.502
21)	B Endrin ke...	7.091	7.897	157.1E6	99005711	54.771	51.790
22)	Mirex	7.274	8.347	109.7E6	70835580	51.896	50.960

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL021521\
Data File : PL064795.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 15 Feb 2021 09:39
Operator : AR\AJ
Sample : PSTDCCC050
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampleId :
PSTDCCC050

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Feb 15 14:41:07 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL020821.M
Quant Title : GC Extractables
QLast Update : Tue Feb 09 04:41:29 2021
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

