

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL021522\
 Data File : PL072712.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 15 Feb 2022 17:10
 Operator : AR\AJ
 Sample : N1513-02
 Misc :
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 20220066-096-21-SITE-WATER

Manual Integrations
 APPROVED

Reviewed By :Abdul Mirza 02/16/2022
 Supervised By :Ankita Jodhani 02/16/2022

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Feb 16 05:14:46 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL020722.M
 Quant Title : GC Extractables
 QLast Update : Tue Feb 08 05:14:10 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.553	5.488	9550452	32776756	19.834	18.975
28) SA Decachlor...	10.308	11.376	15217011	31520805	20.100m	19.877

Target Compounds

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL021522\
Data File : PL072712.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 15 Feb 2022 17:10
Operator : AR\AJ
Sample : N1513-02
Misc :
ALS Vial : 21 Sample Multiplier: 1



Instrument :
ECD_L
ClientSampleId :
20220066-096-21-SITE-WATER


Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 02/16/2022
Supervised By :Ankita Jodhani 02/16/2022

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Feb 16 05:14:46 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL020722.M
Quant Title : GC Extractables
QLast Update : Tue Feb 08 05:14:10 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm






#1 Tetrachloro-m-xylene

R.T.: 4.553 min
Delta R.T.: -0.004 min
Response: 9550452
Conc: 19.83 ng/ml

Instrument :
ECD_L
ClientSampleId :
20220066-096-21-SITE-WATER


Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 02/16/2022
Supervised By :Ankita Jodhani 02/16/2022




#1 Tetrachloro-m-xylene

R.T.: 5.488 min
Delta R.T.: -0.004 min
Response: 32776756
Conc: 18.97 ng/ml



#28 Decachlorobiphenyl

R.T.: 10.308 min
Delta R.T.: -0.009 min
Response: 15217011
Conc: 20.10 ng/ml m



#28 Decachlorobiphenyl

R.T.: 11.376 min
Delta R.T.: -0.010 min
Response: 31520805
Conc: 19.88 ng/ml