

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL021522\
 Data File : PL072732.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 16 Feb 2022 01:25
 Operator : AR\AJ
 Sample : PB142709TB
 Misc :
 ALS Vial : 38 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 PB142709TB

Manual Integrations
 APPROVED

Reviewed By : Abdul Mirza 02/16/2022
 Supervised By : Ankita Jodhani 02/16/2022

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Feb 16 06:00:11 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL020722.M
 Quant Title : GC Extractables
 QLast Update : Tue Feb 08 05:14:10 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.554	5.489	9592763	33325495	19.922	19.292
2) SA Decachlor...	10.309	11.377	17438835	37187545	23.035m	23.451

Target Compounds

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL021522\
Data File : PL072732.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 16 Feb 2022 01:25
Operator : AR\AJ
Sample : PB142709TB
Misc :
ALS Vial : 38 Sample Multiplier: 1



Instrument :
ECD_L
ClientSampleId :
PB142709TB


Manual Integrations
APPROVED

Reviewed By : Abdul Mirza 02/16/2022
Supervised By : Ankita Jodhani 02/16/2022

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Feb 16 06:00:11 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL020722.M
Quant Title : GC Extractables
QLast Update : Tue Feb 08 05:14:10 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm






#1 Tetrachloro-m-xylene

R.T.: 4.554 min
Delta R.T.: -0.003 min
Response: 9592763
Conc: 19.92 ng/ml

Instrument :
ECD_L
ClientSampleId :
PB142709TB


Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 02/16/2022
Supervised By :Ankita Jodhani 02/16/2022




#1 Tetrachloro-m-xylene

R.T.: 5.489 min
Delta R.T.: -0.003 min
Response: 33325495
Conc: 19.29 ng/ml



#28 Decachlorobiphenyl

R.T.: 10.309 min
Delta R.T.: -0.008 min
Response: 17438835
Conc: 23.04 ng/ml m



#28 Decachlorobiphenyl

R.T.: 11.377 min
Delta R.T.: -0.009 min
Response: 37187545
Conc: 23.45 ng/ml