

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL021524\
 Data File : PL088099.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 15 Feb 2024 19:37
 Operator : AR\AJ
 Sample : PSTDICC100
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 PSTDICC100

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Feb 16 11:22:10 2024
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL021524.M
 Quant Title : GC Extractables
 QLast Update : Fri Feb 16 11:18:07 2024
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.635	2.854	281.9E6	143.1E6	101.117	101.175
28)	SA Decachlor...	9.304	8.085	217.1E6	125.6E6	102.262	99.326
Target Compounds							
2)	A alpha-BHC	4.100	3.369	455.9E6	226.7E6	103.056	103.288
3)	MA gamma-BHC...	4.438	3.706	431.4E6	214.6E6	102.208	102.480
4)	MA Heptachlor	5.038	4.054	396.1E6	189.7E6	101.224	101.597
5)	MB Aldrin	5.386	4.340	400.6E6	205.9E6	101.788	102.189
6)	B beta-BHC	4.635	4.007	171.7E6	86808390	99.547	99.643
7)	B delta-BHC	4.887	4.242	410.8E6	216.4E6	102.309	102.851
8)	B Heptachlo...	5.818	4.850	346.2E6	177.9E6	100.564	100.904
9)	A Endosulfan I	6.208	5.227	316.0E6	164.4E6	100.699	100.870
10)	B gamma-Chl...	6.077	5.105	349.0E6	183.8E6	101.224	101.460
11)	B alpha-Chl...	6.158	5.170	345.5E6	181.9E6	100.850	101.011
12)	B 4,4'-DDE	6.332	5.363	322.7E6	171.2E6	102.310	102.632
13)	MA Dieldrin	6.487	5.496	343.4E6	182.6E6	101.885	102.029
14)	MA Endrin	6.720	5.775	276.3E6	154.1E6	102.471	101.395
15)	B Endosulfa...	6.941	6.072	285.9E6	153.3E6	100.586	100.532
16)	A 4,4'-DDD	6.855	5.926	264.8E6	143.2E6	101.902	101.625
17)	MA 4,4'-DDT	7.173	6.180	266.8E6	144.3E6	100.882	102.021
18)	B Endrin al...	7.072	6.254	207.9E6	109.7E6	100.270	100.054
19)	B Endosulfa...	7.309	6.480	269.4E6	143.5E6	101.688	100.745
20)	A Methoxychlor	7.654	6.764	140.7E6	76559463	103.621	99.305
21)	B Endrin ke...	7.801	6.993	298.1E6	170.1E6	102.572	100.664
22)	Mirex	8.282	7.180	216.5E6	128.5E6	96.642	99.399

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL021524\
Data File : PL088099.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 15 Feb 2024 19:37
Operator : AR\AJ
Sample : PSTDICC100
Misc :
ALS Vial : 5 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampleId :
PSTDICC100

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Feb 16 11:22:10 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL021524.M
Quant Title : GC Extractables
QLast Update : Fri Feb 16 11:18:07 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

