

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL021524\
 Data File : PL088100.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 15 Feb 2024 19:51
 Operator : AR\AJ
 Sample : PSTDICC075
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 PSTDICC075

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Feb 16 11:24:03 2024
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL021524.M
 Quant Title : GC Extractables
 QLast Update : Fri Feb 16 11:18:07 2024
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.635	2.854	197.8E6	100.0E6	72.251	72.105
28)	SA Decachlor...	9.304	8.084	120.0E6	89896806	61.590	72.357
Target Compounds							
2)	A alpha-BHC	4.099	3.369	313.2E6	155.4E6	72.142	72.170
3)	MA gamma-BHC...	4.438	3.706	299.2E6	148.3E6	72.213	72.169
4)	MA Heptachlor	5.038	4.054	276.8E6	132.0E6	72.100	72.098
5)	MB Aldrin	5.385	4.340	278.2E6	142.9E6	72.076	72.236
6)	B beta-BHC	4.634	4.007	121.8E6	61747430	72.003	72.200
7)	B delta-BHC	4.887	4.242	284.7E6	149.1E6	72.212	72.189
8)	B Heptachlo...	5.817	4.851	243.8E6	124.4E6	72.157	72.004
9)	A Endosulfan I	6.208	5.226	222.1E6	115.0E6	72.138	71.977
10)	B gamma-Chl...	6.076	5.105	243.8E6	127.8E6	72.072	71.980
11)	B alpha-Chl...	6.157	5.170	242.5E6	126.9E6	72.127	71.923
12)	B 4,4'-DDE	6.331	5.362	223.4E6	117.8E6	72.163	72.026
13)	MA Dieldrin	6.487	5.496	239.5E6	126.2E6	72.331	71.963
14)	MA Endrin	6.719	5.775	192.2E6	107.0E6	72.473	71.865
15)	B Endosulfa...	6.940	6.072	203.2E6	107.5E6	72.632	71.915
16)	A 4,4'-DDD	6.854	5.925	185.8E6	99227739	72.632	71.898
17)	MA 4,4'-DDT	7.173	6.180	187.0E6	100.0E6	72.080	72.100
18)	B Endrin al...	7.071	6.254	148.3E6	77125513	72.662	71.822
19)	B Endosulfa...	7.308	6.480	191.0E6	100.6E6	73.039	72.039
20)	A Methoxychlor	7.653	6.763	100.4E6	54489652	74.313	72.063
21)	B Endrin ke...	7.800	6.993	209.7E6	119.3E6	73.069	72.026
22)	Mirex	8.282	7.179	157.7E6	91320402	71.874	72.037

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL021524\
Data File : PL088100.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 15 Feb 2024 19:51
Operator : AR\AJ
Sample : PSTDICC075
Misc :
ALS Vial : 6 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampleId :
PSTDICC075

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Feb 16 11:24:03 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL021524.M
Quant Title : GC Extractables
QLast Update : Fri Feb 16 11:18:07 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

