

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL021820\
 Data File : PL056651.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 18 Feb 2020 20:43
 Operator : AJ\MA
 Sample : L1544-03
 Misc :
 ALS Vial : 32 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 2-5-1

Manual Integrations
 APPROVED

mohammad
 2/21/2020 3:45:53 PM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Feb 19 00:38:10 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL021220.M
 Quant Title : GC Extractables
 QLast Update : Thu Feb 13 05:32:36 2020
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.497	4.019	204.1E6	64008669	23.131	23.529m
28)	SA Decachlor...	8.251	9.093	233.6E6	50479829	24.384	24.748

Target Compounds

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL021820\
Data File : PL056651.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 18 Feb 2020 20:43
Operator : AJ\MA
Sample : L1544-03
Misc :
ALS Vial : 32 Sample Multiplier: 1



Instrument :
ECD_L
ClientSampleId :
2-5-1


Manual Integrations
APPROVED

mohammad
2/21/2020 3:45:53 PM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Feb 19 00:38:10 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL021220.M
Quant Title : GC Extractables
QLast Update : Thu Feb 13 05:32:36 2020
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm






#1 Tetrachloro-m-xylene

R.T.: 3.497 min
Delta R.T.: 0.000 min
Response: 204101006
Conc: 23.13 ng/ml

Instrument :
ECD_L
ClientSampleId :
2-5-1


Manual Integrations
APPROVED

mohammad
2/21/2020 3:45:53 PM




#1 Tetrachloro-m-xylene

R.T.: 4.019 min
Delta R.T.: 0.001 min
Response: 64008669
Conc: 23.53 ng/ml m



#28 Decachlorobiphenyl

R.T.: 8.251 min
Delta R.T.: 0.002 min
Response: 233583732
Conc: 24.38 ng/ml



#28 Decachlorobiphenyl

R.T.: 9.093 min
Delta R.T.: 0.003 min
Response: 50479829
Conc: 24.75 ng/ml