

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL021821\
 Data File : PL064827.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 18 Feb 2021 18:19
 Operator : AR\AJ
 Sample : PSTDCCC050
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 PSTDCCC050

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Feb 19 03:25:54 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL020821.M
 Quant Title : GC Extractables
 QLast Update : Tue Feb 09 04:41:29 2021
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.424	4.099	144.7E6	120.0E6	51.446	55.530
28)	SA Decachlor...	8.144	9.255	106.8E6	87095020	47.474	45.854
Target Compounds							
2)	A alpha-BHC	3.857	4.502	247.0E6	178.7E6	53.262	58.988
3)	MA gamma-BHC...	4.138	4.793	224.4E6	164.0E6	51.908	57.732
4)	MA Heptachlor	4.434	5.304	202.0E6	152.7E6	49.255	51.533
5)	MB Aldrin	4.679	5.614	193.4E6	135.4E6	51.949	55.203
6)	B beta-BHC	4.388	4.972	88283880	75615760	51.653	58.387
7)	B delta-BHC	4.590	5.193	217.8E6	147.4E6	53.877	54.569
8)	B Heptachlo...	5.120	6.009	179.8E6	124.6E6	51.793	56.473
9)	A Endosulfan I	5.455	6.371	163.7E6	112.3E6	51.034	53.532
10)	B gamma-Chl...	5.345	6.244	180.2E6	123.7E6	51.379	53.520
11)	B alpha-Chl...	5.403	6.317	176.4E6	123.1E6	50.924	53.491
12)	B 4,4'-DDE	5.569	6.474	159.8E6	108.1E6	51.053	53.297
13)	MA Dieldrin	5.697	6.632	168.1E6	113.6E6	51.233	53.161
14)	MA Endrin	5.951	6.853	135.4E6	84769517	45.096	46.891
15)	B Endosulfa...	6.223	7.068	148.2E6	96613691	51.461	51.272
16)	A 4,4'-DDD	6.081	6.973	142.2E6	93901449	56.142	54.166
17)	MA 4,4'-DDT	6.320	7.272	99412951	68707418	44.086	45.064
18)	B Endrin al...	6.392	7.194	123.3E6	84203392	50.571	51.710
19)	B Endosulfa...	6.605	7.418	143.5E6	93708106	50.607	52.071
20)	A Methoxychlor	6.867	7.733	54250453	40117315	41.897	44.380
21)	B Endrin ke...	7.091	7.896	154.8E6	99657884	53.971	52.131
22)	Mirex	7.273	8.346	105.6E6	69881661	49.961	50.274

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL021821\
Data File : PL064827.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 18 Feb 2021 18:19
Operator : AR\AJ
Sample : PSTDCCC050
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampleId :
PSTDCCC050

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Feb 19 03:25:54 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL020821.M
Quant Title : GC Extractables
QLast Update : Tue Feb 09 04:41:29 2021
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

