

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL021821\
 Data File : PL064813.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 18 Feb 2021 14:59
 Operator : AR\AJ
 Sample : M1422-01
 Misc :
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 BU-03-021621

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Feb 18 17:02:00 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL020821.M
 Quant Title : GC Extractables
 QLast Update : Tue Feb 09 04:41:29 2021
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.425	4.099	62637767	51174695	22.263	23.690
28)	SA Decachlor...	8.147	9.255	47068473	35524639	20.930	18.703
Target Compounds							
10)	B gamma-Chl...	5.346	6.244	81891341	67934101	23.355	29.396 #
11)	B alpha-Chl...	5.403	6.316	88453095	88199896	25.538	38.311 #
12)	B 4,4'-DDE	5.569	6.473	24342792	17285134	7.779	8.523
17)	MA 4,4'-DDT	6.321	7.269	3435159	2363463	1.523	1.550m



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL021821\
Data File : PL064813.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 18 Feb 2021 14:59
Operator : AR\AJ
Sample : M1422-01
Misc :
ALS Vial : 7 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampled :
BU-03-021621

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Feb 18 17:02:00 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL020821.M
Quant Title : GC Extractables
QLast Update : Tue Feb 09 04:41:29 2021
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm






#1 Tetrachloro-m-xylene


R.T.: 3.425 min
Delta R.T.: 0.000 min
Response: 62637767
Conc: 22.26 ng/ml

Instrument :
ECD_L
ClientSampleId :
BU-03-021621




#1 Tetrachloro-m-xylene

R.T.: 4.099 min
Delta R.T.: -0.003 min
Response: 51174695
Conc: 23.69 ng/ml




#10 gamma-Chlordane

R.T.: 5.346 min
Delta R.T.: -0.003 min
Response: 81891341
Conc: 23.35 ng/ml



#10 gamma-Chlordane


R.T.: 6.244 min
Delta R.T.: -0.005 min
Response: 67934101
Conc: 29.40 ng/ml



#11 alpha-Chlordane


R.T.: 5.403 min
Delta R.T.: -0.003 min
Response: 88453095
Conc: 25.54 ng/ml

Instrument :
ECD_L
ClientSampleId :
BU-03-021621




#11 alpha-Chlordane

R.T.: 6.316 min
Delta R.T.: -0.005 min
Response: 88199896
Conc: 38.31 ng/ml




#12 4,4'-DDE

R.T.: 5.569 min
Delta R.T.: -0.002 min
Response: 24342792
Conc: 7.78 ng/ml



#12 4,4'-DDE


R.T.: 6.473 min
Delta R.T.: -0.005 min
Response: 17285134
Conc: 8.52 ng/ml



#17 4,4'-DDT


R.T.: 6.321 min
Delta R.T.: -0.002 min
Response: 3435159
Conc: 1.52 ng/ml

Instrument :
ECD_L
ClientSampleId :
BU-03-021621




#17 4,4'-DDT

R.T.: 7.269 min
Delta R.T.: -0.007 min
Response: 2363463
Conc: 1.55 ng/ml m



#28 Decachlorobiphenyl

R.T.: 8.147 min
Delta R.T.: 0.000 min
Response: 47068473
Conc: 20.93 ng/ml



#28 Decachlorobiphenyl

R.T.: 9.255 min
Delta R.T.: -0.006 min
Response: 35524639
Conc: 18.70 ng/ml