

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL022020\
 Data File : PL056700.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 20 Feb 2020 15:59
 Operator : AJ\MA
 Sample : PSTDCCC050
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 PSTDCCC050

Manual Integrations
 APPROVED

mohammad
 2/25/2020 2:44:40 PM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Feb 21 00:29:55 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL021920.M
 Quant Title : GC Extractables
 QLast Update : Thu Feb 20 02:32:45 2020
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.498	4.019	462.4E6	127.9E6	53.471	51.358
28)	SA Decachlor...	8.251	9.090	560.9E6	104.5E6	47.153	49.744
Target Compounds							
2)	A alpha-BHC	3.936	4.408	805.8E6	197.6E6	51.052	52.533
3)	MA gamma-BHC...	4.220	4.693	767.6E6	181.2E6	51.456	52.621
4)	MA Heptachlor	4.521	5.200	754.2E6	154.2E6	51.147	52.693
5)	MB Aldrin	4.770	5.505	761.7E6	157.7E6	51.018	50.382
6)	B beta-BHC	4.470	4.866	301.7E6	72145946	52.053	55.184
7)	B delta-BHC	4.675	5.083	756.5E6	163.3E6	52.485	53.647
8)	B Heptachlo...	5.212	5.892	710.4E6	157.2E6	50.213	53.771
9)	A Endosulfan I	5.551	6.250	670.7E6	136.4E6	50.169m	50.474
10)	B gamma-Chl...	5.439	6.127	757.1E6	145.3E6	52.120	50.644
11)	B alpha-Chl...	5.497	6.199	746.7E6	143.9E6	51.711	50.563
12)	B 4,4'-DDE	5.662	6.358	695.7E6	136.2E6	49.461	49.457
13)	MA Dieldrin	5.794	6.507	718.0E6	144.9E6	48.172m	50.675
14)	MA Endrin	6.053	6.725	683.2E6	120.2E6	50.619	48.799
15)	B Endosulfa...	6.326	6.936	675.4E6	130.7E6	54.387	50.615
16)	A 4,4'-DDD	6.178	6.849	628.0E6	112.1E6	56.148	48.674m
17)	MA 4,4'-DDT	6.417	7.148	583.5E6	107.0E6	53.582	51.430
18)	B Endrin al...	6.494	7.060	543.3E6	104.8E6	50.188	49.963
19)	B Endosulfa...	6.707	7.284	594.0E6	119.1E6	50.101	50.130
20)	A Methoxychlor	6.962	7.608	283.7E6	58397751	53.505	52.767
21)	B Endrin ke...	7.199	7.755	668.1E6	127.3E6	50.116	50.406
22)	Mirex	7.387	8.208	446.5E6	85986953	53.203	51.184

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL022020\
Data File : PL056700.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 20 Feb 2020 15:59
Operator : AJ\MA
Sample : PSTDCCC050
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampleId :
PSTDCCC050

Manual Integrations
APPROVED

mohammad
2/25/2020 2:44:40 PM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Feb 21 00:29:55 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL021920.M
Quant Title : GC Extractables
QLast Update : Thu Feb 20 02:32:45 2020
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

