

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL022023\
 Data File : PL081024.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 20 Feb 2023 13:28
 Operator : AR\AJ
 Sample : PSTDICC100
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 PSTDICC100

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Feb 20 14:38:32 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL022023.M
 Quant Title : GC Extractables
 QLast Update : Mon Feb 20 14:37:41 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.352	2.688	236.2E6	234.4E6	101.139	100.839
28)	SA Decachlor...	8.792	7.835	184.3E6	221.9E6	94.861	96.213
Target Compounds							
2)	A alpha-BHC	3.801	3.189	383.3E6	365.9E6	106.016	105.913
3)	MA gamma-BHC...	4.129	3.517	352.7E6	338.6E6	104.853	105.069
4)	MA Heptachlor	4.713	3.858	339.0E6	347.8E6	102.314	103.455
5)	MB Aldrin	5.052	4.137	313.6E6	329.0E6	103.531	105.046
6)	B beta-BHC	4.326	3.814	144.5E6	135.7E6	97.436	99.194
7)	B delta-BHC	4.570	4.043	342.4E6	330.6E6	104.496	106.266
8)	B Heptachlo...	5.477	4.639	281.2E6	295.8E6	100.364	102.275
9)	A Endosulfan I	5.859	5.009	261.8E6	276.6E6	100.781	102.685
10)	B gamma-Chl...	5.732	4.891	286.8E6	305.8E6	102.144	103.963
11)	B alpha-Chl...	5.811	4.955	282.2E6	302.8E6	101.328	102.847
12)	B 4,4'-DDE	5.989	5.147	261.9E6	260.1E6	104.397	106.934
13)	MA Dieldrin	6.134	5.275	274.9E6	294.2E6	102.824	104.989
14)	MA Endrin	6.361	5.549	244.7E6	262.2E6	102.383	104.193
15)	B Endosulfa...	6.580	5.843	221.1E6	235.4E6	100.278	102.348
16)	A 4,4'-DDD	6.503	5.701	211.1E6	206.2E6	103.089	106.418
17)	MA 4,4'-DDT	6.817	5.952	232.6E6	234.2E6	101.947	105.850
18)	B Endrin al...	6.709	6.023	178.4E6	185.4E6	98.198	100.109
19)	B Endosulfa...	6.943	6.246	227.1E6	240.3E6	98.568	100.984
20)	A Methoxychlor	7.295	6.531	117.8E6	118.6E6	96.126	98.643
21)	B Endrin ke...	7.423	6.751	238.6E6	267.8E6	99.477	101.324
22)	Mirex	7.895	6.934	175.1E6	221.3E6	95.118	96.690

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL022023\
Data File : PL081024.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 20 Feb 2023 13:28
Operator : AR\AJ
Sample : PSTDICC100
Misc :
ALS Vial : 5 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampleId :
PSTDICC100

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Feb 20 14:38:32 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL022023.M
Quant Title : GC Extractables
QLast Update : Mon Feb 20 14:37:41 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

