

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL022120\
 Data File : PL056739.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 21 Feb 2020 17:22
 Operator : AJ\MA
 Sample : PB127007BS
 Misc :
 ALS Vial : 32 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 PB127007BS

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Feb 21 22:27:08 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL021920.M
 Quant Title : GC Extractables
 QLast Update : Thu Feb 20 02:32:45 2020
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.498	4.019	167.4E6	51860027	19.359	20.819
2)	SA Decachlor...	8.251	9.089	200.1E6	45287801	16.823	21.552 #
Target Compounds							
2)	A alpha-BHC	3.936	4.408	721.2E6	191.1E6	45.694	50.811
3)	MA gamma-BHC...	4.220	4.693	674.3E6	173.0E6	45.201	50.243
4)	MA Heptachlor	4.521	5.199	683.3E6	148.5E6	46.338	50.759
5)	MB Aldrin	4.770	5.504	686.8E6	154.5E6	46.003	49.372
6)	B beta-BHC	4.471	4.865	281.0E6	74670521	48.491	57.115
7)	B delta-BHC	4.676	5.082	656.1E6	153.2E6	45.521	50.336
8)	B Heptachlo...	5.213	5.892	652.9E6	143.3E6	46.154	49.020
9)	A Endosulfan I	5.553	6.249	624.0E6	135.6E6	46.677	50.200
10)	B gamma-Chl...	5.440	6.125	674.3E6	143.3E6	46.421	49.970
11)	B alpha-Chl...	5.498	6.198	650.9E6	140.8E6	45.079	49.478
12)	B 4,4'-DDE	5.662	6.358	629.8E6	138.1E6	44.774	50.143
13)	MA Dieldrin	5.796	6.506	680.7E6	144.2E6	45.671	50.448
14)	MA Endrin	6.054	6.724	588.8E6	121.2E6	43.627	49.216
15)	B Endosulfa...	6.327	6.935	554.0E6	122.2E6	44.609	47.318
16)	A 4,4'-DDD	6.179	6.849	482.8E6	111.8E6	43.159	48.544
17)	MA 4,4'-DDT	6.418	7.147	467.6E6	104.4E6	42.933	50.182
18)	B Endrin al...	6.494	7.059	493.3E6	105.5E6	45.569	50.270
19)	B Endosulfa...	6.707	7.283	568.6E6	121.1E6	47.953	50.970
20)	A Methoxychlor	6.962	7.607	249.1E6	56886563	46.982	51.401
21)	B Endrin ke...	7.199	7.754	644.8E6	132.0E6	48.369	52.277
22)	Mirex	7.388	8.207	442.7E6	90685763	52.760	53.981

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL022120\
Data File : PL056739.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 21 Feb 2020 17:22
Operator : AJ\MA
Sample : PB127007BS
Misc :
ALS Vial : 32 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampled :
PB127007BS

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Feb 21 22:27:08 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL021920.M
Quant Title : GC Extractables
QLast Update : Thu Feb 20 02:32:45 2020
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

