

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL022121\
 Data File : PL064845.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 21 Feb 2021 13:51
 Operator : AR\AJ
 Sample : INDB101
 Misc :
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Feb 22 03:17:21 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL022121CLP.M
 Quant Title : GC Extractables
 QLast Update : Mon Feb 22 03:16:29 2021
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml


System Monitoring Compounds							
1)	SA Tetrachlo...	3.423	4.098	14944074	12068368	5.117	5.002
27)	SA Decachlor...	8.142	9.252	30972034	23584353	10.813	10.460
Target Compounds							
5)	MB Aldrin	4.677	5.612	18274617	13043286	4.762	4.805
6)	B beta-BHC	4.386	4.971	9431258	12005071	5.043	7.002 #
7)	B delta-BHC	4.587	5.192	20877248	14307449	4.959	5.254
8)	B Heptachlo...	5.118	6.007	17919073	9931287	4.923	4.119
10)	B trans-Chl...	5.343	6.243	18538484	13188822	4.940	4.933
11)	B cis-Chlor...	5.400	6.315	18001113	13202478	4.936	4.995
12)	B 4,4'-DDE	5.566	6.472	32074308	22078401	9.253	9.244
15)	B Endosulfa...	6.222	7.066	29108429	19756985	9.793	9.874
18)	B Endrin al...	6.390	7.192	25781569	17909648	10.191	10.190
19)	B Endosulfa...	6.602	7.416	32214416	20936782	10.539	10.315
21)	B Endrin ke...	7.088	7.894	30161857	19974124	9.646	9.761


(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL022121\
Data File : PL064845.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 21 Feb 2021 13:51
Operator : AR\AJ
Sample : INDB101
Misc :
ALS Vial : 10 Sample Multiplier: 1

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Feb 22 03:17:21 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL022121CLP.M
Quant Title : GC Extractables
QLast Update : Mon Feb 22 03:16:29 2021
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm






#1 Tetrachloro-m-xylene


R.T.: 3.423 min
Delta R.T.: 0.000 min
Response: 14944074
Conc: 5.12 ng/ml

Instrument :
ECD_L
ClientSampleId :




#1 Tetrachloro-m-xylene

R.T.: 4.098 min
Delta R.T.: 0.000 min
Response: 12068368
Conc: 5.00 ng/ml




#5 Aldrin

R.T.: 4.677 min
Delta R.T.: 0.000 min
Response: 18274617
Conc: 4.76 ng/ml



#5 Aldrin


R.T.: 5.612 min
Delta R.T.: 0.000 min
Response: 13043286
Conc: 4.80 ng/ml



#6 beta-BHC


R.T.: 4.386 min
Delta R.T.: 0.000 min
Response: 9431258
Conc: 5.04 ng/ml

Instrument :
ECD_L
ClientSampleId :




#6 beta-BHC

R.T.: 4.971 min
Delta R.T.: 0.000 min
Response: 12005071
Conc: 7.00 ng/ml




#7 delta-BHC

R.T.: 4.587 min
Delta R.T.: 0.000 min
Response: 20877248
Conc: 4.96 ng/ml



#7 delta-BHC


R.T.: 5.192 min
Delta R.T.: 0.000 min
Response: 14307449
Conc: 5.25 ng/ml



#8 Heptachlor epoxide


R.T.: 5.118 min
Delta R.T.: 0.000 min
Response: 17919073
Conc: 4.92 ng/ml

Instrument :
ECD_L
ClientSampleId :




#8 Heptachlor epoxide

R.T.: 6.007 min
Delta R.T.: 0.000 min
Response: 9931287
Conc: 4.12 ng/ml




#10 trans-Chlordane

R.T.: 5.343 min
Delta R.T.: 0.000 min
Response: 18538484
Conc: 4.94 ng/ml



#10 trans-Chlordane


R.T.: 6.243 min
Delta R.T.: 0.000 min
Response: 13188822
Conc: 4.93 ng/ml



#11 cis-Chlordane


R.T.: 5.400 min
Delta R.T.: 0.000 min
Response: 18001113
Conc: 4.94 ng/ml

Instrument :
ECD_L
ClientSampleId :




#11 cis-Chlordane

R.T.: 6.315 min
Delta R.T.: 0.000 min
Response: 13202478
Conc: 5.00 ng/ml




#12 4,4'-DDE

R.T.: 5.566 min
Delta R.T.: 0.000 min
Response: 32074308
Conc: 9.25 ng/ml



#12 4,4'-DDE


R.T.: 6.472 min
Delta R.T.: 0.000 min
Response: 22078401
Conc: 9.24 ng/ml



#15 Endosulfan II


R.T.: 6.222 min
Delta R.T.: 0.000 min
Response: 29108429
Conc: 9.79 ng/ml

Instrument :
ECD_L
ClientSampleId :




#15 Endosulfan II

R.T.: 7.066 min
Delta R.T.: 0.000 min
Response: 19756985
Conc: 9.87 ng/ml




#18 Endrin aldehyde

R.T.: 6.390 min
Delta R.T.: 0.000 min
Response: 25781569
Conc: 10.19 ng/ml



#18 Endrin aldehyde


R.T.: 7.192 min
Delta R.T.: 0.000 min
Response: 17909648
Conc: 10.19 ng/ml



#19 Endosulfan Sulfate


R.T.: 6.602 min
Delta R.T.: 0.000 min
Response: 32214416
Conc: 10.54 ng/ml

Instrument :
ECD_L
ClientSampleId :




#19 Endosulfan Sulfate

R.T.: 7.416 min
Delta R.T.: 0.000 min
Response: 20936782
Conc: 10.32 ng/ml




#21 Endrin ketone

R.T.: 7.088 min
Delta R.T.: 0.000 min
Response: 30161857
Conc: 9.65 ng/ml



#21 Endrin ketone


R.T.: 7.894 min
Delta R.T.: 0.000 min
Response: 19974124
Conc: 9.76 ng/ml



#27 Decachlorobiphenyl

R.T.: 8.142 min
Delta R.T.: 0.000 min
Response: 30972034
Conc: 10.81 ng/ml

Instrument :
ECD_L
ClientSampleId :



#27 Decachlorobiphenyl

R.T.: 9.252 min
Delta R.T.: 0.000 min
Response: 23584353
Conc: 10.46 ng/ml