

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL022121\
 Data File : PL064853.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 21 Feb 2021 15:45
 Operator : AR\AJ
 Sample : INDB501
 Misc :
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Feb 22 03:18:03 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL022121CLP.M
 Quant Title : GC Extractables
 QLast Update : Mon Feb 22 03:16:29 2021
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml


System Monitoring Compounds							
1)	SA Tetrachlo...	3.424	4.098	226.5E6	188.4E6	77.553	78.063
27)	SA Decachlor...	8.142	9.252	404.5E6	319.8E6	141.240	141.858
Target Compounds							
5)	MB Aldrin	4.678	5.613	312.0E6	226.4E6	81.309	83.398
6)	B beta-BHC	4.387	4.971	138.8E6	116.0E6	74.218	67.644
7)	B delta-BHC	4.589	5.192	338.3E6	226.2E6	80.360	83.051
8)	B Heptachlo...	5.119	6.007	282.9E6	195.6E6	77.715	81.120
10)	B trans-Chl...	5.344	6.244	294.9E6	211.0E6	78.576	78.914
11)	B cis-Chlor...	5.401	6.316	282.6E6	205.0E6	77.499	77.553
12)	B 4,4'-DDE	5.568	6.472	566.9E6	404.9E6	163.539	169.549
15)	B Endosulfa...	6.223	7.067	457.8E6	317.3E6	154.003	158.568
18)	B Endrin al...	6.391	7.192	375.8E6	267.9E6	148.556	152.444
19)	B Endosulfa...	6.603	7.416	457.2E6	309.5E6	149.576	152.469
21)	B Endrin ke...	7.089	7.895	488.2E6	327.2E6	156.137	159.902


(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL022121\
Data File : PL064853.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 21 Feb 2021 15:45
Operator : AR\AJ
Sample : INDB501
Misc :
ALS Vial : 18 Sample Multiplier: 1

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Feb 22 03:18:03 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL022121CLP.M
Quant Title : GC Extractables
QLast Update : Mon Feb 22 03:16:29 2021
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm






#1 Tetrachloro-m-xylene


R.T.: 3.424 min
Delta R.T.: 0.000 min
Response: 226486962
Conc: 77.55 ng/ml

Instrument :
ECD_L
ClientSampleId :




#1 Tetrachloro-m-xylene

R.T.: 4.098 min
Delta R.T.: 0.000 min
Response: 188360697
Conc: 78.06 ng/ml




#5 Aldrin

R.T.: 4.678 min
Delta R.T.: 0.000 min
Response: 312005158
Conc: 81.31 ng/ml



#5 Aldrin


R.T.: 5.613 min
Delta R.T.: 0.000 min
Response: 226387084
Conc: 83.40 ng/ml



#6 beta-BHC


R.T.: 4.387 min
Delta R.T.: 0.000 min
Response: 138791205
Conc: 74.22 ng/ml

Instrument :
ECD_L
ClientSampleId :




#6 beta-BHC

R.T.: 4.971 min
Delta R.T.: 0.000 min
Response: 115969597
Conc: 67.64 ng/ml




#7 delta-BHC

R.T.: 4.589 min
Delta R.T.: 0.000 min
Response: 338300349
Conc: 80.36 ng/ml



#7 delta-BHC


R.T.: 5.192 min
Delta R.T.: 0.000 min
Response: 226155174
Conc: 83.05 ng/ml



#8 Heptachlor epoxide


R.T.: 5.119 min
Delta R.T.: 0.000 min
Response: 282884437
Conc: 77.71 ng/ml

Instrument :
ECD_L
ClientSampleId :




#8 Heptachlor epoxide

R.T.: 6.007 min
Delta R.T.: 0.000 min
Response: 195597378
Conc: 81.12 ng/ml




#10 trans-Chlordane

R.T.: 5.344 min
Delta R.T.: 0.001 min
Response: 294877637
Conc: 78.58 ng/ml



#10 trans-Chlordane


R.T.: 6.244 min
Delta R.T.: 0.000 min
Response: 210994382
Conc: 78.91 ng/ml



#11 cis-Chlordane


R.T.: 5.401 min
Delta R.T.: 0.000 min
Response: 282631210
Conc: 77.50 ng/ml

Instrument :
ECD_L
ClientSampleId :




#11 cis-Chlordane

R.T.: 6.316 min
Delta R.T.: 0.000 min
Response: 204965953
Conc: 77.55 ng/ml




#12 4,4'-DDE

R.T.: 5.568 min
Delta R.T.: 0.000 min
Response: 566893416
Conc: 163.54 ng/ml



#12 4,4'-DDE


R.T.: 6.472 min
Delta R.T.: 0.000 min
Response: 404948476
Conc: 169.55 ng/ml



#15 Endosulfan II


R.T.: 6.223 min
Delta R.T.: 0.000 min
Response: 457768845
Conc: 154.00 ng/ml

Instrument :
ECD_L
ClientSampleId :




#15 Endosulfan II

R.T.: 7.067 min
Delta R.T.: 0.000 min
Response: 317275006
Conc: 158.57 ng/ml




#18 Endrin aldehyde

R.T.: 6.391 min
Delta R.T.: 0.001 min
Response: 375807115
Conc: 148.56 ng/ml



#18 Endrin aldehyde


R.T.: 7.192 min
Delta R.T.: 0.000 min
Response: 267925888
Conc: 152.44 ng/ml



#19 Endosulfan Sulfate


R.T.: 6.603 min
Delta R.T.: 0.000 min
Response: 457197018
Conc: 149.58 ng/ml

Instrument :
ECD_L
ClientSampleId :




#19 Endosulfan Sulfate

R.T.: 7.416 min
Delta R.T.: 0.000 min
Response: 309473274
Conc: 152.47 ng/ml




#21 Endrin ketone

R.T.: 7.089 min
Delta R.T.: 0.000 min
Response: 488235639
Conc: 156.14 ng/ml



#21 Endrin ketone


R.T.: 7.895 min
Delta R.T.: 0.000 min
Response: 327214950
Conc: 159.90 ng/ml



#27 Decachlorobiphenyl

R.T.: 8.142 min
Delta R.T.: 0.000 min
Response: 404542584
Conc: 141.24 ng/ml

Instrument :
ECD_L
ClientSampleId :



#27 Decachlorobiphenyl

R.T.: 9.252 min
Delta R.T.: 0.000 min
Response: 319835094
Conc: 141.86 ng/ml