

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL022222\
 Data File : PL072879.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 22 Feb 2022 11:37
 Operator : AR\AJ
 Sample : N1579-06DL 20X
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 SB-4R-N(0-0.5)DL

Manual Integrations
 APPROVED

Reviewed By : Abdul Mirza 02/23/2022
 Supervised By : Ankita Jodhani 02/23/2022

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Feb 23 05:31:28 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL020722.M
 Quant Title : GC Extractables
 QLast Update : Tue Feb 08 05:14:10 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.556	5.488	478063	1455971	0.993	0.843
28) SA Decachlor...	10.314	11.380	1003042	1650901	1.325	1.041
Target Compounds						
12) B 4,4'-DDE	7.553	8.491	24398786	71157722	40.149	38.071
17) MA 4,4'-DDT	8.399	9.345	25113779	67567980	46.673	44.138m

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL022222\
Data File : PL072879.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 22 Feb 2022 11:37
Operator : AR\AJ
Sample : N1579-06DL 20X
Misc :
ALS Vial : 5 Sample Multiplier: 1



Instrument :
ECD_L
ClientSampleId :
SB-4R-N(0-0.5)DL


Manual Integrations
APPROVED

Reviewed By : Abdul Mirza 02/23/2022
Supervised By : Ankita Jodhani 02/23/2022

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Feb 23 05:31:28 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL020722.M
Quant Title : GC Extractables
QLast Update : Tue Feb 08 05:14:10 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm






#1 Tetrachloro-m-xylene

R.T.: 4.556 min
Delta R.T.: -0.001 min
Response: 478063
Conc: 0.99 ng/ml

Instrument :
ECD_L
ClientSampleId :
SB-4R-N(0-0.5)DL


Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 02/23/2022
Supervised By :Ankita Jodhani 02/23/2022




#1 Tetrachloro-m-xylene

R.T.: 5.488 min
Delta R.T.: -0.004 min
Response: 1455971
Conc: 0.84 ng/ml




#12 4,4'-DDE

R.T.: 7.553 min
Delta R.T.: -0.002 min
Response: 24398786
Conc: 40.15 ng/ml



#12 4,4'-DDE

R.T.: 8.491 min
Delta R.T.: -0.003 min
Response: 71157722
Conc: 38.07 ng/ml




#17 4,4'-DDT

R.T.: 8.399 min
Delta R.T.: -0.002 min
Response: 25113779
Conc: 46.67 ng/ml

Instrument :
ECD_L
ClientSampleId :
SB-4R-N(0-0.5)DL


Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 02/23/2022
Supervised By :Ankita Jodhani 02/23/2022




#17 4,4'-DDT

R.T.: 9.345 min
Delta R.T.: -0.006 min
Response: 67567980
Conc: 44.14 ng/ml m



#28 Decachlorobiphenyl

R.T.: 10.314 min
Delta R.T.: -0.003 min
Response: 1003042
Conc: 1.32 ng/ml



#28 Decachlorobiphenyl

R.T.: 11.380 min
Delta R.T.: -0.006 min
Response: 1650901
Conc: 1.04 ng/ml