

Data Path : P:\HPCHEM1\ECD_L\Data\PL022318\
 Data File : PL033385.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 24 Feb 2018 06:09
 Operator : UA\AJ
 Sample : J1648-01
 Misc :
 ALS Vial : 56 Sample Multiplier: 1

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Feb 26 03:56:34 2018
 Quant Method : P:\HPCHEM1\ECD_L\methods\PL022318.M
 Quant Title : GC Extractables
 QLast Update : Fri Feb 23 13:36:10 2018
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml



System Monitoring Compounds							
1)	SA Tetrachlo...	3.402	3.985	71087095	20853085	12.065	12.900
28)	SA Decachlor...	8.104	9.011	80172618	19106798	10.304	12.386
Target Compounds							
2)	A alpha-BHC	3.793f	0.000	3719119	0	0.403	N.D. #
6)	B beta-BHC	4.330f	0.000	9705593	0	2.587	N.D. #
8)	B Heptachlo...	5.145f	0.000	4417886	0	0.552	N.D. #
10)	B gamma-Chl...	5.292	0.000	4666566	0	0.554	N.D. #
11)	B alpha-Chl...	5.334f	0.000	7921275	0	0.959	N.D. #
12)	B 4,4'-DDE	5.538	0.000	13356223	0	1.598	N.D. #
13)	MA Dieldrin	5.663	0.000	17125286	0	2.042	N.D. #
14)	MA Endrin	5.979f	0.000	10224630	0	1.275	N.D. #
17)	MA 4,4'-DDT	6.285	0.000	12726517	0	1.750	N.D. #
25)	Chlordane-3	5.292	0.000	4666566	0	4.672	N.D. #
26)	Chlordane-4	5.334	0.000	7921275	0	8.801	N.D. #
27)	Chlordane-5	5.979	0.000	10224630	0	65.711	N.D. #


(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : P:\HPCHEM1\ECD_L\Data\PL022318\
Data File : PL033385.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 24 Feb 2018 06:09
Operator : UA\AJ
Sample : J1648-01
Misc :
ALS Vial : 56 Sample Multiplier: 1

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Feb 26 03:56:34 2018
Quant Method : P:\HPCHEM1\ECD_L\methods\PL022318.M
Quant Title : GC Extractables
QLast Update : Fri Feb 23 13:36:10 2018
Response via : Initial Calibration
Integrator: ChemStation


Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x0.25µm






#1 Tetrachloro-m-xylene

R.T.: 3.402 min
Delta R.T.: 0.000 min
Response: 71087095
Conc: 12.07 ng/ml




#1 Tetrachloro-m-xylene

R.T.: 3.985 min
Delta R.T.: 0.000 min
Response: 20853085
Conc: 12.90 ng/ml




#2 alpha-BHC

R.T.: 3.793 min
Delta R.T.: -0.046 min
Response: 3719119
Conc: 0.40 ng/ml




#2 alpha-BHC

R.T.: 0.000 min
Exp R.T.: 4.364 min
Response: 0
Conc: N.D.




#6 beta-BHC

R.T.: 4.330 min
Delta R.T.: -0.041 min
Response: 9705593
Conc: 2.59 ng/ml




#6 beta-BHC

R.T.: 0.000 min
Exp R.T.: 4.805 min
Response: 0
Conc: N.D.




#8 Heptachlor epoxide

R.T.: 5.145 min
Delta R.T.: 0.055 min
Response: 4417886
Conc: 0.55 ng/ml




#8 Heptachlor epoxide

R.T.: 0.000 min
Exp R.T.: 5.840 min
Response: 0
Conc: N.D.




#10 gamma-Chlordane

R.T.: 5.292 min
Delta R.T.: -0.021 min
Response: 4666566
Conc: 0.55 ng/ml




#10 gamma-Chlordane

R.T.: 0.000 min
Exp R.T.: 6.073 min
Response: 0
Conc: N.D.




#11 alpha-Chlordane

R.T.: 5.334 min
Delta R.T.: -0.036 min
Response: 7921275
Conc: 0.96 ng/ml




#11 alpha-Chlordane

R.T.: 0.000 min
Exp R.T.: 6.147 min
Response: 0
Conc: N.D.




#12 4,4'-DDE

R.T.: 5.538 min
Delta R.T.: 0.001 min
Response: 13356223
Conc: 1.60 ng/ml




#12 4,4'-DDE

R.T.: 0.000 min
Exp R.T. : 6.302 min
Response: 0
Conc: N.D.




#13 Dieldrin

R.T.: 5.663 min
Delta R.T.: 0.000 min
Response: 17125286
Conc: 2.04 ng/ml




#13 Dieldrin

R.T.: 0.000 min
Exp R.T. : 6.446 min
Response: 0
Conc: N.D.




#14 Endrin

R.T.: 5.979 min
Delta R.T.: 0.059 min
Response: 10224630
Conc: 1.28 ng/ml




#14 Endrin

R.T.: 0.000 min
Exp R.T.: 6.660 min
Response: 0
Conc: N.D.




#17 4,4'-DDT

R.T.: 6.285 min
Delta R.T.: -0.001 min
Response: 12726517
Conc: 1.75 ng/ml




#17 4,4'-DDT

R.T.: 0.000 min
Exp R.T.: 7.082 min
Response: 0
Conc: N.D.




#25 Chlordane-3

R.T.: 5.292 min
Delta R.T.: -0.020 min
Response: 4666566
Conc: 4.67 ng/ml




#25 Chlordane-3

R.T.: 0.000 min
Exp R.T.: 6.074 min
Response: 0
Conc: N.D.




#26 Chlordane-4

R.T.: 5.334 min
Delta R.T.: -0.036 min
Response: 7921275
Conc: 8.80 ng/ml




#26 Chlordane-4

R.T.: 0.000 min
Exp R.T.: 6.150 min
Response: 0
Conc: N.D.




#27 Chlordane-5

R.T.: 5.979 min
Delta R.T.: -0.001 min
Response: 10224630
Conc: 65.71 ng/ml




#27 Chlordane-5

R.T.: 0.000 min
Exp R.T.: 6.940 min
Response: 0
Conc: N.D.



#28 Decachlorobiphenyl

R.T.: 8.104 min
Delta R.T.: 0.003 min
Response: 80172618
Conc: 10.30 ng/ml



#28 Decachlorobiphenyl

R.T.: 9.011 min
Delta R.T.: 0.003 min
Response: 19106798
Conc: 12.39 ng/ml