

Method Path : P:\HPCHEM1\ECD_L\methods\
 Method File : LTX022318.M
 Title : GC Extractables
 Last Update : Fri Feb 23 12:19:29 2018
 Response Via : Initial Calibration

Calibration Files

500 =PL033312.D 1000 =PL033310.D 250 =PL033313.D
 750 =PL033311.D =

	Compound	500	1000	250	750	Avg	%RSD
1) SA	Tetrachloro-m-xylene	5.572	5.558	5.388	5.563	5.520 E6	1.60
2)	Toxaphene-1	4.150	4.223	4.063	4.253	4.173 E4	2.03
3)	Toxaphene-2	1.269	1.287	1.221	1.285	1.266 E5	2.43
4)	Toxaphene-3	1.080	1.059	1.026	1.038	1.051 E5	2.27
5)	Toxaphene-4	5.319	5.016	4.940	4.838	5.028 E4	4.12
6)	Toxaphene-5	1.440	1.456	1.346	1.538	1.445 E5	5.44
7) SA	Decachlorobiphenyl	7.102	6.912	7.155	6.978	7.037 E6	1.59

Signal #2 Calibration Files

500 =PL033312.D 1000 =PL033310.D 250 =PL033313.D
 750 =PL033311.D =

	Compound	500	1000	250	750	Avg	%RSD
1) SA	Tetrachloro-m-xylene	1.534	1.516	1.528	1.524	1.525 E6	0.49
2)	Toxaphene-1	9.221	8.455	9.623	9.010	9.077 E3	5.36
3)	Toxaphene-2	1.523	1.440	1.525	1.470	1.490 E4	2.79
4)	Toxaphene-3	4.124	3.918	4.143	3.963	4.037 E4	2.80
5)	Toxaphene-4	2.369	2.227	2.293	2.299	2.297 E4	2.52
6)	Toxaphene-5	1.992	1.855	2.097	1.903	1.962 E4	5.42
7) SA	Decachlorobiphenyl	1.459	1.371	1.520	1.407	1.439 E6	4.49

(#) = Out of Range