

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL022324\
 Data File : PL088270.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 23 Feb 2024 19:12
 Operator : AR\AJ
 Sample : P1564-15
 Misc :
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 LKWD-BACKFILL-16

Manual Integrations
 APPROVED

Reviewed By : Abdul Mirza 02/26/2024
 Supervised By : Ankita Jodhani 02/26/2024

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Feb 24 00:25:51 2024
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL021524.M
 Quant Title : GC Extractables
 QLast Update : Fri Feb 16 21:46:54 2024
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.633	2.853	57042367	30126416	20.767m	21.480
28) SA Decachlor...	9.307	8.087	39421734	24736722	20.469	19.427

Target Compounds

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL022324\
Data File : PL088270.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 23 Feb 2024 19:12
Operator : AR\AJ
Sample : P1564-15
Misc :
ALS Vial : 24 Sample Multiplier: 1



Instrument :
ECD_L
ClientSampleId :
LKWD-BACKFILL-16


Manual Integrations
APPROVED

Reviewed By : Abdul Mirza 02/26/2024
Supervised By : Ankita Jodhani 02/26/2024

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Feb 24 00:25:51 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL021524.M
Quant Title : GC Extractables
QLast Update : Fri Feb 16 21:46:54 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm






#1 Tetrachloro-m-xylene

R.T.: 3.633 min
Delta R.T.: -0.002 min
Response: 57042367
Conc: 20.77 ng/ml

Instrument :
ECD_L
ClientSampleId :
LKWD-BACKFILL-16


Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 02/26/2024
Supervised By :Ankita Jodhani 02/26/2024




#1 Tetrachloro-m-xylene

R.T.: 2.853 min
Delta R.T.: 0.000 min
Response: 30126416
Conc: 21.48 ng/ml



#28 Decachlorobiphenyl

R.T.: 9.307 min
Delta R.T.: 0.003 min
Response: 39421734
Conc: 20.47 ng/ml



#28 Decachlorobiphenyl

R.T.: 8.087 min
Delta R.T.: 0.002 min
Response: 24736722
Conc: 19.43 ng/ml