

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL022420\
 Data File : PL056806.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 24 Feb 2020 17:54
 Operator : AJ\MA
 Sample : PSTDCCC050
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 PSTDCCC050

Manual Integrations
 APPROVED

mohammad
 2/26/2020 12:32:01 PM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Feb 25 02:39:59 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL021920.M
 Quant Title : GC Extractables
 QLast Update : Thu Feb 20 02:32:45 2020
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.498	4.020	426.3E6	129.1E6	49.294	51.831
28)	SA Decachlor...	8.252	9.093	510.8E6	110.5E6	42.943	52.599
Target Compounds							
2)	A alpha-BHC	3.936	4.409	750.0E6	201.1E6	47.518	53.462
3)	MA gamma-BHC...	4.220	4.694	710.6E6	185.8E6	47.632	53.959
4)	MA Heptachlor	4.522	5.201	675.3E6	151.0E6	45.797	51.618
5)	MB Aldrin	4.771	5.505	701.4E6	161.4E6	46.979	51.585
6)	B beta-BHC	4.471	4.867	280.2E6	74023767	48.352	56.620
7)	B delta-BHC	4.676	5.084	702.9E6	162.8E6	48.768	53.482
8)	B Heptachlo...	5.213	5.893	659.4E6	143.6E6	46.614	49.105
9)	A Endosulfan I	5.554	6.251	628.5E6	137.0E6	47.014	50.710
10)	B gamma-Chl...	5.440	6.128	683.8E6	146.5E6	47.072	51.067
11)	B alpha-Chl...	5.499	6.200	670.9E6	145.2E6	46.458	51.017
12)	B 4,4'-DDE	5.661	6.358	653.0E6	139.3E6	46.423m	50.595
13)	MA Dieldrin	5.795	6.508	679.1E6	144.9E6	45.560m	50.680
14)	MA Endrin	6.054	6.726	586.1E6	113.7E6	43.422	46.159
15)	B Endosulfa...	6.327	6.937	576.2E6	124.8E6	46.392	48.323
16)	A 4,4'-DDD	6.179	6.851	548.4E6	118.9E6	49.028	51.619
17)	MA 4,4'-DDT	6.420	7.148	459.1E6	96541170	42.155	46.403
18)	B Endrin al...	6.496	7.061	490.0E6	103.4E6	45.267	49.300
19)	B Endosulfa...	6.709	7.285	550.8E6	116.9E6	46.454	49.171
20)	A Methoxychlor	6.964	7.609	236.2E6	51548895	44.554	46.578
21)	B Endrin ke...	7.200	7.756	632.7E6	128.4E6	47.459	50.839
22)	Mirex	7.390	8.209	408.4E6	83939614	48.664	49.965

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL022420\
Data File : PL056806.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 24 Feb 2020 17:54
Operator : AJ\MA
Sample : PSTDCCC050
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampleId :
PSTDCCC050

Manual Integrations
APPROVED

mohammad
2/26/2020 12:32:01 PM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Feb 25 02:39:59 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL021920.M
Quant Title : GC Extractables
QLast Update : Thu Feb 20 02:32:45 2020
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

