

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL022423\
 Data File : PL081128.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 24 Feb 2023 16:26
 Operator : AR\AJ
 Sample : PSTDCCC050
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 PSTDCCC050

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Feb 25 00:30:18 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL022023.M
 Quant Title : GC Extractables
 QLast Update : Mon Feb 20 18:06:39 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.351	2.686	118.5E6	114.2E6	50.822	49.615
28)	SA Decachlor...	8.790	7.830	93498777	111.3E6	47.823	47.529
Target Compounds							
2)	A alpha-BHC	3.799	3.186	182.1E6	165.7E6	51.679	49.480
3)	MA gamma-BHC...	4.128	3.515	169.5E6	155.5E6	51.467	49.520
4)	MA Heptachlor	4.711	3.855	162.7E6	157.5E6	49.465	48.045
5)	MB Aldrin	5.050	4.134	150.9E6	149.4E6	50.166	48.604
6)	B beta-BHC	4.324	3.811	74160913	66630635	49.621	49.524
7)	B delta-BHC	4.567	4.040	165.4E6	150.1E6	51.507	49.688
8)	B Heptachlo...	5.475	4.636	139.6E6	136.4E6	50.147	47.681
9)	A Endosulfan I	5.857	5.005	128.8E6	127.9E6	49.642	47.986
10)	B gamma-Chl...	5.730	4.887	140.5E6	139.1E6	50.312	47.776
11)	B alpha-Chl...	5.809	4.951	138.7E6	139.4E6	49.998	47.738
12)	B 4,4'-DDE	5.987	5.143	123.9E6	119.7E6	50.064	50.136
13)	MA Dieldrin	6.132	5.270	131.7E6	132.8E6	49.746	48.224
14)	MA Endrin	6.358	5.545	111.8E6	111.3E6	47.233	45.022
15)	B Endosulfa...	6.577	5.838	108.5E6	108.7E6	48.897	48.089
16)	A 4,4'-DDD	6.501	5.697	104.3E6	100.5E6	51.542	52.781
17)	MA 4,4'-DDT	6.814	5.948	104.8E6	101.6E6	46.384	46.853
18)	B Endrin al...	6.707	6.019	87810740	87040905	48.291	47.173
19)	B Endosulfa...	6.941	6.242	113.2E6	113.4E6	49.139	47.841
20)	A Methoxychlor	7.292	6.526	55470323	55705745	45.205	46.411
21)	B Endrin ke...	7.421	6.747	118.9E6	128.3E6	49.909	49.009
22)	Mirex	7.892	6.929	90323936	107.5E6	48.278	46.233

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL022423\
Data File : PL081128.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 24 Feb 2023 16:26
Operator : AR\AJ
Sample : PSTDCCC050
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampleId :
PSTDCCC050

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Feb 25 00:30:18 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL022023.M
Quant Title : GC Extractables
QLast Update : Mon Feb 20 18:06:39 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

