

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL022522\
 Data File : PL072979.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 25 Feb 2022 21:52
 Operator : AR\AJ
 Sample : INDA501
 Misc :
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 INDA5324

Manual Integrations
 APPROVED

Reviewed By : Abdul Mirza 03/01/2022
 Supervised By : Ankita Jodhani 03/02/2022

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Mar 01 01:27:22 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL022522CLP.M
 Quant Title : GC Extractables
 QLast Update : Tue Mar 01 01:24:23 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml


System Monitoring Compounds							
1)	SA Tetrachlo...	4.646	5.502	26019092	120.8E6	77.805	78.593m
27)	SA Decachlor...	10.427	11.337	69125148	232.5E6	150.528	150.820m
Target Compounds							
2)	A alpha-BHC	5.356	6.066	35917424	169.5E6	83.850	83.992
3)	MA gamma-BHC...	5.779	6.454	36480133	159.3E6	82.470	82.061m
4)	MA Heptachlor	6.192	7.112	36705121	161.1E6	79.394	79.664m
9)	A Endosulfan I	7.498	8.337	34536530	124.3E6	78.781	78.962
13)	MA Dieldrin	7.785	8.621	77976536	274.5E6	163.884	162.701m
14)	MA Endrin	8.076	8.854	67059964	218.9E6	162.890	162.248m
16)	A 4,4'-DDD	8.243	8.997	60169284	208.3E6	165.605	160.653m
17)	MA 4,4'-DDT	8.505	9.315	65292602	216.7E6	164.017	162.335m
20)	A Methoxychlor	9.101	9.795	185.4E6	588.3E6	776.882	784.280

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL022522\
Data File : PL072979.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 25 Feb 2022 21:52
Operator : AR\AJ
Sample : INDA501
Misc :
ALS Vial : 17 Sample Multiplier: 1

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 01 01:27:22 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL022522CLP.M
Quant Title : GC Extractables
QLast Update : Tue Mar 01 01:24:23 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm



Reviewed By: Abdul Mirza 03/01/2022
Supervised By: Ankita Jodhani 03/02/2022

Manual Integrations
APPROVED

Instrument :
ECD_L
Client Sampled :
INDAS324