

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL022620\
 Data File : PL056854.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 26 Feb 2020 22:17
 Operator : AJ\MA
 Sample : PSTDCCC050
 Misc :
 ALS Vial : 100 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 PSTDCCC050

Manual Integrations
 APPROVED

mohammad
 2/27/2020 11:38:40 AM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Feb 27 00:37:33 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL021920.M
 Quant Title : GC Extractables
 QLast Update : Thu Feb 20 02:32:45 2020
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.498	4.021	439.7E6	129.6E6	50.851	52.046
28)	SA Decachlor...	8.253	9.096	482.4E6	102.2E6	40.551	48.653
Target Compounds							
2)	A alpha-BHC	3.937	4.410	769.4E6	201.7E6	48.747	53.607
3)	MA gamma-BHC...	4.221	4.695	728.5E6	187.6E6	48.833	54.464
4)	MA Heptachlor	4.522	5.202	720.7E6	159.7E6	48.877	54.573
5)	MB Aldrin	4.771	5.507	717.3E6	163.6E6	48.047	52.287
6)	B beta-BHC	4.472	4.868	284.5E6	75097928	49.089	57.442
7)	B delta-BHC	4.677	5.085	713.5E6	160.4E6	49.506	52.672
8)	B Heptachlo...	5.215	5.895	675.4E6	151.6E6	47.743	51.853
9)	A Endosulfan I	5.554	6.252	670.6E6	138.6E6	50.166	51.295
10)	B gamma-Chl...	5.441	6.129	703.2E6	148.4E6	48.409	51.734
11)	B alpha-Chl...	5.500	6.202	694.6E6	147.3E6	48.102	51.769
12)	B 4,4'-DDE	5.662	6.361	636.1E6	141.3E6	45.226m	51.310
13)	MA Dieldrin	5.796	6.510	691.6E6	146.8E6	46.400m	51.353
14)	MA Endrin	6.055	6.728	633.8E6	122.6E6	46.960	49.765
15)	B Endosulfa...	6.329	6.938	598.1E6	126.0E6	48.158	48.776
16)	A 4,4'-DDD	6.180	6.852	546.1E6	116.8E6	48.821	50.717
17)	MA 4,4'-DDT	6.421	7.150	523.8E6	108.0E6	48.096	51.913
18)	B Endrin al...	6.497	7.063	496.5E6	104.9E6	45.870	50.017
19)	B Endosulfa...	6.710	7.287	549.6E6	117.7E6	46.354	49.514
20)	A Methoxychlor	6.965	7.611	258.1E6	58145572	48.683	52.539
21)	B Endrin ke...	7.202	7.758	604.9E6	126.3E6	45.373	50.000
22)	Mirex	7.391	8.211	386.2E6	83296231	46.028	49.582

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL022620\
Data File : PL056854.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 26 Feb 2020 22:17
Operator : AJ\MA
Sample : PSTDCCC050
Misc :
ALS Vial : 100 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampleId :
PSTDCCC050

Manual Integrations
APPROVED

mohammad
2/27/2020 11:38:40 AM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Feb 27 00:37:33 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL021920.M
Quant Title : GC Extractables
QLast Update : Thu Feb 20 02:32:45 2020
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

