

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL022720\
 Data File : PL056858.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 27 Feb 2020 11:19
 Operator : AJ\MA
 Sample : RESCHK
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 RESCHK

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Feb 27 13:53:10 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL022720.M
 Quant Title : GC Extractables
 QLast Update : Thu Feb 27 13:35:45 2020
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.498	4.020	406.9E6	121.7E6	22.753	22.704
28)	SA Decachlor...	8.254	9.098	457.0E6	95363092	23.009	22.532
Target Compounds							
9)	A Endosulfan I	5.555	6.252	267.3E6	61352068	10.742	10.934
10)	B gamma-Chl...	5.442	6.130	284.4E6	65213526	10.409	10.881
12)	B 4,4'-DDE	5.664	6.362	559.4E6	122.9E6	21.735	21.117
13)	MA Dieldrin	5.798	6.510	587.4E6	127.1E6	21.673	21.346
19)	B Endosulfa...	6.710	7.288	450.2E6	97494083	21.140	20.181
20)	A Methoxychlor	6.966	7.612	1068.9E6	232.7E6	103.174	99.504
21)	B Endrin ke...	7.202	7.760	510.0E6	105.8E6	20.706	20.557

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL022720\
Data File : PL056858.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 27 Feb 2020 11:19
Operator : AJ\MA
Sample : RESCHK
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampled :
RESCHK

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Feb 27 13:53:10 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL022720.M
Quant Title : GC Extractables
QLast Update : Thu Feb 27 13:35:45 2020
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

