

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL022720\
 Data File : PL056861.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 27 Feb 2020 12:01
 Operator : AJ\MA
 Sample : PSTDICC050
 Misc :
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 PSTDICC050

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Feb 27 13:15:48 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL022720.M
 Quant Title : GC Extractables
 QLast Update : Thu Feb 27 13:13:46 2020
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.498	4.020	931.3E6	274.1E6	50.000	50.000
28)	SA Decachlor...	8.255	9.097	1008.2E6	208.5E6	50.000	50.000
Target Compounds							
2)	A alpha-BHC	3.936	4.410	1615.0E6	428.6E6	50.000	50.000
3)	MA gamma-BHC...	4.220	4.695	1534.8E6	396.4E6	50.000	50.000
4)	MA Heptachlor	4.522	5.202	1476.7E6	327.9E6	50.000	50.000
5)	MB Aldrin	4.771	5.508	1483.9E6	339.0E6	50.000	50.000
6)	B beta-BHC	4.471	4.868	592.2E6	163.5E6	50.000	50.000
7)	B delta-BHC	4.677	5.085	1497.6E6	332.8E6	50.000	50.000
8)	B Heptachlo...	5.214	5.895	1376.6E6	302.7E6	50.000	50.000
9)	A Endosulfan I	5.554	6.253	1308.9E6	282.8E6	50.000	50.000
10)	B gamma-Chl...	5.442	6.130	1430.5E6	303.4E6	50.000	50.000
11)	B alpha-Chl...	5.500	6.202	1398.7E6	300.1E6	50.000	50.000
12)	B 4,4'-DDE	5.664	6.361	1349.4E6	295.2E6	50.000	50.000
13)	MA Dieldrin	5.798	6.510	1423.1E6	302.7E6	50.000	50.000
14)	MA Endrin	6.056	6.729	1268.9E6	259.4E6	50.000	50.000
15)	B Endosulfa...	6.329	6.939	1172.2E6	265.7E6	50.000	50.000
16)	A 4,4'-DDD	6.181	6.853	1047.2E6	241.9E6	50.000	50.000
17)	MA 4,4'-DDT	6.421	7.152	1082.0E6	231.4E6	50.000	50.000
18)	B Endrin al...	6.497	7.064	1027.6E6	219.3E6	50.000	50.000
19)	B Endosulfa...	6.710	7.287	1132.7E6	239.7E6	50.000	50.000
20)	A Methoxychlor	6.965	7.611	532.7E6	116.4E6	50.000	50.000
21)	B Endrin ke...	7.202	7.759	1287.5E6	258.6E6	50.000	50.000
22)	Mirex	7.391	8.212	792.6E6	166.0E6	50.000	50.000

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL022720\
Data File : PL056861.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 27 Feb 2020 12:01
Operator : AJ\MA
Sample : PSTDICC050
Misc :
ALS Vial : 7 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampleId :
PSTDICC050

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Feb 27 13:15:48 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL022720.M
Quant Title : GC Extractables
QLast Update : Thu Feb 27 13:13:46 2020
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

