

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL022819\
 Data File : PL045216.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 28 Feb 2019 08:56
 Operator : AJ\SJ
 Sample : PEM
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 PEM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Mar 01 00:00:59 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL022719.M
 Quant Title : GC Extractables
 QLast Update : Wed Feb 27 17:00:03 2019
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.364	3.983	149.9E6	46815810	21.797	21.861
28)	SA Decachlor...	8.048	9.040	189.7E6	53117762	22.666	23.247
Target Compounds							
2)	A alpha-BHC	3.794	4.373	106.1E6	29612394	9.550	9.745
3)	MA gamma-BHC...	4.071	4.657	101.4E6	29434112	10.030	10.421
6)	B beta-BHC	4.324	4.829	48251636	14507640	11.258	11.424
8)	B Heptachlo...	0.000	5.850	0	5563725	N.D.	2.316 #
14)	MA Endrin	5.863	6.686	497.3E6	118.6E6	55.223	52.120
16)	A 4,4'-DDD	5.996	6.809	8611074	2514607	1.176	1.308
17)	MA 4,4'-DDT	6.232	7.108	919.5E6	226.1E6	122.603	114.572
20)	A Methoxychlor	6.777	7.565	1204.1E6	311.9E6	274.170	254.136
21)	B Endrin ke...	6.999	7.711	367619	3217471	0.040	1.353 #



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL022819\
Data File : PL045216.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 28 Feb 2019 08:56
Operator : AJ\SJ
Sample : PEM
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampleId :
PEM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 01 00:00:59 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL022719.M
Quant Title : GC Extractables
QLast Update : Wed Feb 27 17:00:03 2019
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x0.25µm






#1 Tetrachloro-m-xylene


R.T.: 3.364 min
Delta R.T.: 0.000 min
Response: 149944852
Conc: 21.80 ng/ml

Instrument :
ECD_L
ClientSampleId :
PEM




#1 Tetrachloro-m-xylene

R.T.: 3.983 min
Delta R.T.: 0.000 min
Response: 46815810
Conc: 21.86 ng/ml




#2 alpha-BHC

R.T.: 3.794 min
Delta R.T.: 0.000 min
Response: 106070344
Conc: 9.55 ng/ml



#2 alpha-BHC


R.T.: 4.373 min
Delta R.T.: 0.001 min
Response: 29612394
Conc: 9.75 ng/ml



#3 gamma-BHC (Lindane)


R.T.: 4.071 min
Delta R.T.: 0.000 min
Response: 101373174
Conc: 10.03 ng/ml

Instrument :
ECD_L
ClientSampleId :
PEM




#3 gamma-BHC (Lindane)

R.T.: 4.657 min
Delta R.T.: 0.000 min
Response: 29434112
Conc: 10.42 ng/ml




#6 beta-BHC

R.T.: 4.324 min
Delta R.T.: 0.000 min
Response: 48251636
Conc: 11.26 ng/ml



#6 beta-BHC


R.T.: 4.829 min
Delta R.T.: 0.002 min
Response: 14507640
Conc: 11.42 ng/ml



#8 Heptachlor epoxide


R.T.: 0.000 min
Exp R.T.: 5.037 min
Response: 0
Conc: N.D.

Instrument :
ECD_L
ClientSampleId :
PEM




#8 Heptachlor epoxide

R.T.: 5.850 min
Delta R.T.: -0.004 min
Response: 5563725
Conc: 2.32 ng/ml




#14 Endrin

R.T.: 5.863 min
Delta R.T.: 0.000 min
Response: 497322505
Conc: 55.22 ng/ml



#14 Endrin


R.T.: 6.686 min
Delta R.T.: 0.001 min
Response: 118587392
Conc: 52.12 ng/ml



#16 4,4'-DDD


R.T.: 5.996 min
Delta R.T.: 0.000 min
Response: 8611074
Conc: 1.18 ng/ml

Instrument :
ECD_L
ClientSampleId :
PEM




#16 4,4'-DDD

R.T.: 6.809 min
Delta R.T.: 0.001 min
Response: 2514607
Conc: 1.31 ng/ml




#17 4,4'-DDT

R.T.: 6.232 min
Delta R.T.: 0.000 min
Response: 919539225
Conc: 122.60 ng/ml



#17 4,4'-DDT


R.T.: 7.108 min
Delta R.T.: 0.002 min
Response: 226102635
Conc: 114.57 ng/ml



#20 Methoxychlor


R.T.: 6.777 min
Delta R.T.: 0.000 min
Response: 1204078628
Conc: 274.17 ng/ml

Instrument :
ECD_L
ClientSampleId :
PEM




#20 Methoxychlor

R.T.: 7.565 min
Delta R.T.: 0.000 min
Response: 311930663
Conc: 254.14 ng/ml




#21 Endrin ketone

R.T.: 6.999 min
Delta R.T.: 0.000 min
Response: 367619
Conc: 0.04 ng/ml



#21 Endrin ketone


R.T.: 7.711 min
Delta R.T.: 0.001 min
Response: 3217471
Conc: 1.35 ng/ml



#28 Decachlorobiphenyl

R.T.: 8.048 min
Delta R.T.: 0.000 min
Response: 189659479
Conc: 22.67 ng/ml

Instrument :
ECD_L
ClientSampleId :
PEM



#28 Decachlorobiphenyl

R.T.: 9.040 min
Delta R.T.: 0.001 min
Response: 53117762
Conc: 23.25 ng/ml