

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL022819\
 Data File : PL045234.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 28 Feb 2019 13:18
 Operator : AJ\SJ
 Sample : K1627-07
 Misc :
 ALS Vial : 75 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 SB-04-(0-2)

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Mar 01 00:05:44 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL022719.M
 Quant Title : GC Extractables
 QLast Update : Wed Feb 27 17:00:03 2019
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.363	3.982	50346701	16583798	7.319	7.744
28)	SA Decachlor...	8.047	9.040	65593243	17005432	7.839	7.442
Target Compounds							
10)	B gamma-Chl...	5.281f	0.000	16264330	0	1.662	N.D. #
11)	B alpha-Chl...	0.000	6.163	0	6974925	N.D.	2.586 #
12)	B 4,4'-DDE	5.485	6.318	80609773	23977070	9.183	10.265
16)	A 4,4'-DDD	0.000	6.814	0	2318172	N.D.	1.205 #
17)	MA 4,4'-DDT	6.230	7.106	22353693	3463797	2.980	1.755 #
20)	A Methoxychlor	6.755f	0.000	4683516	0	1.066	N.D. #
25)	Chlordane-3	5.281f	0.000	16264330	0	14.180	N.D. #
26)	Chlordane-4	0.000	6.163	0	6974925	N.D.	14.909 #



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL022819\
Data File : PL045234.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 28 Feb 2019 13:18
Operator : AJ\SJ
Sample : K1627-07
Misc :
ALS Vial : 75 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampleId :
SB-04-(0-2)

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 01 00:05:44 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL022719.M
Quant Title : GC Extractables
QLast Update : Wed Feb 27 17:00:03 2019
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x0.25µm






#1 Tetrachloro-m-xylene


R.T.: 3.363 min
Delta R.T.: 0.000 min
Response: 50346701
Conc: 7.32 ng/ml

Instrument :
ECD_L
ClientSampleId :
SB-04-(0-2)




#1 Tetrachloro-m-xylene

R.T.: 3.982 min
Delta R.T.: 0.000 min
Response: 16583798
Conc: 7.74 ng/ml




#10 gamma-Chlordane

R.T.: 5.281 min
Delta R.T.: 0.020 min
Response: 16264330
Conc: 1.66 ng/ml



#10 gamma-Chlordane


R.T.: 0.000 min
Exp R.T. : 6.088 min
Response: 0
Conc: N.D.



#11 alpha-Chlordane


R.T.: 0.000 min
Exp R.T.: 5.317 min
Response: 0
Conc: N.D.

Instrument :
ECD_L
ClientSampleId :
SB-04-(0-2)




#11 alpha-Chlordane

R.T.: 6.163 min
Delta R.T.: 0.003 min
Response: 6974925
Conc: 2.59 ng/ml




#12 4,4'-DDE

R.T.: 5.485 min
Delta R.T.: 0.000 min
Response: 80609773
Conc: 9.18 ng/ml



#12 4,4'-DDE


R.T.: 6.318 min
Delta R.T.: 0.000 min
Response: 23977070
Conc: 10.26 ng/ml



#16 4,4'-DDD


R.T.: 0.000 min
Exp R.T.: 5.996 min
Response: 0
Conc: N.D.

Instrument :
ECD_L
ClientSampleId :
SB-04-(0-2)




#16 4,4'-DDD

R.T.: 6.814 min
Delta R.T.: 0.006 min
Response: 2318172
Conc: 1.21 ng/ml




#17 4,4'-DDT

R.T.: 6.230 min
Delta R.T.: -0.001 min
Response: 22353693
Conc: 2.98 ng/ml



#17 4,4'-DDT


R.T.: 7.106 min
Delta R.T.: 0.000 min
Response: 3463797
Conc: 1.76 ng/ml



#20 Methoxychlor


R.T.: 6.755 min
Delta R.T.: -0.021 min
Response: 4683516
Conc: 1.07 ng/ml

Instrument :
ECD_L
ClientSampleId :
SB-04-(0-2)




#20 Methoxychlor

R.T.: 0.000 min
Exp R.T. : 7.564 min
Response: 0
Conc: N.D.




#25 Chlordane-3

R.T.: 5.281 min
Delta R.T.: 0.019 min
Response: 16264330
Conc: 14.18 ng/ml



#25 Chlordane-3


R.T.: 0.000 min
Exp R.T. : 6.089 min
Response: 0
Conc: N.D.



#26 Chlordane-4


R.T.: 0.000 min
Exp R.T.: 5.317 min
Response: 0
Conc: N.D.

Instrument :
ECD_L
ClientSampleId :
SB-04-(0-2)




#26 Chlordane-4

R.T.: 6.163 min
Delta R.T.: 0.000 min
Response: 6974925
Conc: 14.91 ng/ml



#28 Decachlorobiphenyl

R.T.: 8.047 min
Delta R.T.: 0.000 min
Response: 65593243
Conc: 7.84 ng/ml



#28 Decachlorobiphenyl

R.T.: 9.040 min
Delta R.T.: 0.000 min
Response: 17005432
Conc: 7.44 ng/ml