

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL022819\
 Data File : PL045217.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 28 Feb 2019 09:17
 Operator : AJ\SJ
 Sample : PSTDCCC050
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 PSTDCCC050

Manual Integrations
 APPROVED

Sohil
 3/1/2019 11:48:38 AM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Mar 01 00:01:12 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL022719.M
 Quant Title : GC Extractables
 QLast Update : Wed Feb 27 17:00:03 2019
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.369	3.982	353.9E6	109.2E6	51.438	51.011
28)	SA Decachlor...	8.054	9.043	420.0E6	115.2E6	50.189	50.409
Target Compounds							
2)	A alpha-BHC	3.799	4.373	583.2E6	159.6E6	52.507	52.529
3)	MA gamma-BHC...	4.077	4.658	530.3E6	148.3E6	52.468	52.485
4)	MA Heptachlor	4.365	5.165	554.9E6	152.6E6	52.582	51.293
5)	MB Aldrin	4.604	5.470	526.7E6	139.7E6	54.156m	51.741
6)	B beta-BHC	4.330	4.830	219.6E6	70297833	51.241	55.355
7)	B delta-BHC	4.526	5.046	521.8E6	124.4E6	54.589	57.839
8)	B Heptachlo...	5.043	5.856	492.9E6	125.3E6	52.808	52.187
9)	A Endosulfan I	5.376	6.213	472.7E6	125.8E6	52.882	50.509
10)	B gamma-Chl...	5.267	6.090	515.7E6	136.0E6	52.693	50.685
11)	B alpha-Chl...	5.323	6.163	508.7E6	135.5E6	52.458	50.223
12)	B 4,4'-DDE	5.491	6.321	472.4E6	119.6E6	53.820	51.185
13)	MA Dieldrin	5.615	6.470	514.2E6	131.0E6	53.801	50.982
14)	MA Endrin	5.870	6.687	464.0E6	111.3E6	51.523	48.916
15)	B Endosulfa...	6.144	6.897	452.6E6	118.4E6	52.215	50.366
16)	A 4,4'-DDD	6.003	6.811	396.8E6	99918765	54.192	51.958
17)	MA 4,4'-DDT	6.238	7.109	398.6E6	101.0E6	53.140	51.161
18)	B Endrin al...	6.311	7.021	378.0E6	102.4E6	51.713	50.085
19)	B Endosulfa...	6.522	7.244	421.2E6	113.0E6	51.822	50.224
20)	A Methoxychlor	6.783	7.567	220.4E6	60840525	50.194	49.568
21)	B Endrin ke...	7.006	7.714	493.8E6	123.6E6	53.596	51.968
22)	Mirex	7.183	8.166	369.7E6	99843175	49.692	50.130

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL022819\
 Data File : PL045217.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 28 Feb 2019 09:17
 Operator : AJ\SJ
 Sample : PSTDCCC050
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampled :
 PSTDCCC050

Manual Integrations
 APPROVED

Sohil
 3/1/2019 11:48:38 AM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Mar 01 00:01:12 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL022719.M
 Quant Title : GC Extractables
 QLast Update : Wed Feb 27 17:00:03 2019
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

