

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL030122\
 Data File : PL073070.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 01 Mar 2022 12:47
 Operator : AR\AJ
 Sample : PSTDCCC050
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 PSTDCCC050

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Mar 01 14:35:10 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL022622.M
 Quant Title : GC Extractables
 QLast Update : Mon Feb 28 02:28:04 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.648	5.490	17535773	83160274	54.721	57.305
2)	SA Decachlor...	10.428	11.329	25311984	86577798	54.077	54.882
Target Compounds							
2)	A alpha-BHC	5.357	6.053	23447402	117.4E6	55.641	59.527
3)	MA gamma-BHC...	5.780	6.442	23435822	108.6E6	52.999	58.066
4)	MA Heptachlor	6.193	7.101	24231669	109.4E6	52.792	54.742
5)	MB Aldrin	6.518	7.469	24415850	106.5E6	55.053	58.199
6)	B beta-BHC	6.160	6.678	11709080	49309591	52.577	54.680
7)	B delta-BHC	6.422	6.946	24156029	108.7E6	55.316	57.450
8)	B Heptachlo...	7.096	7.925	24291346	97729976	52.139	55.521
9)	A Endosulfan I	7.499	8.325	22780926	86750331	51.360	53.553
10)	B gamma-Chl...	7.374	8.193	24834732	97729680	52.296	54.729
11)	B alpha-Chl...	7.443	8.276	24945177	95965391	52.120	54.509
12)	B 4,4'-DDE	7.658	8.460	20964998	86147402	53.938	55.098
13)	MA Dieldrin	7.786	8.610	23688849	90785731	52.864	56.211
14)	MA Endrin	8.076	8.845	21499254	71044295	54.145	53.669
15)	B Endosulfa...	8.387	9.068	23597190	80595183	54.405	58.239
16)	A 4,4'-DDD	8.243	8.987	19023007	73123534	54.620	54.986
17)	MA 4,4'-DDT	8.505	9.305	19704050	69820152	51.778	50.902
18)	B Endrin al...	8.575	9.200	15965527	53265562	49.731	48.949
19)	B Endosulfa...	8.807	9.437	23463952	75013070	53.378	52.908
20)	A Methoxychlor	9.101	9.784	13133916	40600825	52.585	51.860
21)	B Endrin ke...	9.326	9.926	27498157	81937931	55.466	55.884
22)	Mirex	9.515	10.403	24460130	65935213	43.352	44.898

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL030122\
Data File : PL073070.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 01 Mar 2022 12:47
Operator : AR\AJ
Sample : PSTDCCC050
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampleId :
PSTDCCC050

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 01 14:35:10 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL022622.M
Quant Title : GC Extractables
QLast Update : Mon Feb 28 02:28:04 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

