

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL030122\
 Data File : PL073067.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 01 Mar 2022 11:55
 Operator : AR\AJ
 Sample : N1681-01RE
 Misc :
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 TP-1-1-8FTRE

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Mar 01 13:07:23 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL022622.M
 Quant Title : GC Extractables
 QLast Update : Mon Feb 28 02:28:04 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.646	5.490	4222684	18809499	13.177	12.961
28)	SA Decachlor...	10.427	11.328	332.3E6	1153.6E6	709.963	731.251
Target Compounds							
11)	B alpha-Chl...	7.441	8.280	217010	1647799	0.453	0.936 #
12)	B 4,4'-DDE	7.657	8.459	6220612	24633068	16.004	15.755
16)	A 4,4'-DDD	8.242	8.989	664241	4861060	1.907	3.655m#
17)	MA 4,4'-DDT	8.504	9.305	9512394	34380198	24.997	25.064



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL030122\
Data File : PL073067.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 01 Mar 2022 11:55
Operator : AR\AJ
Sample : N1681-01RE
Misc :
ALS Vial : 8 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampleId :
TP-1-1-8FTRE

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 01 13:07:23 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL022622.M
Quant Title : GC Extractables
QLast Update : Mon Feb 28 02:28:04 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm






#1 Tetrachloro-m-xylene


R.T.: 4.646 min
Delta R.T.: -0.002 min
Response: 4222684
Conc: 13.18 ng/ml

Instrument :
ECD_L
ClientSampleId :
TP-1-1-8FTRE




#1 Tetrachloro-m-xylene

R.T.: 5.490 min
Delta R.T.: 0.000 min
Response: 18809499
Conc: 12.96 ng/ml




#11 alpha-Chlordane

R.T.: 7.441 min
Delta R.T.: -0.002 min
Response: 217010
Conc: 0.45 ng/ml



#11 alpha-Chlordane


R.T.: 8.280 min
Delta R.T.: 0.006 min
Response: 1647799
Conc: 0.94 ng/ml



#12 4,4'-DDE


R.T.: 7.657 min
Delta R.T.: 0.000 min
Response: 6220612
Conc: 16.00 ng/ml

Instrument :
ECD_L
ClientSampleId :
TP-1-1-8FTRE




#12 4,4'-DDE

R.T.: 8.459 min
Delta R.T.: 0.000 min
Response: 24633068
Conc: 15.75 ng/ml




#16 4,4'-DDD

R.T.: 8.242 min
Delta R.T.: 0.000 min
Response: 664241
Conc: 1.91 ng/ml



#16 4,4'-DDD


R.T.: 8.989 min
Delta R.T.: 0.004 min
Response: 4861060
Conc: 3.66 ng/ml m



#17 4,4'-DDT


R.T.: 8.504 min
Delta R.T.: 0.000 min
Response: 9512394
Conc: 25.00 ng/ml

Instrument :
ECD_L
ClientSampleId :
TP-1-1-8FTRE




#17 4,4'-DDT

R.T.: 9.305 min
Delta R.T.: 0.001 min
Response: 34380198
Conc: 25.06 ng/ml



#28 Decachlorobiphenyl

R.T.: 10.427 min
Delta R.T.: 0.000 min
Response: 332311687
Conc: 709.96 ng/ml



#28 Decachlorobiphenyl

R.T.: 11.328 min
Delta R.T.: 0.002 min
Response: 1153576861
Conc: 731.25 ng/ml