

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL030223\
 Data File : PL081226.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 03 Mar 2023 04:06
 Operator : AR\AJ
 Sample : 01749-16
 Misc :
 ALS Vial : 47 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 CF-18

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Mar 03 04:50:45 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL030223.M
 Quant Title : GC Extractables
 QLast Update : Fri Mar 03 01:48:45 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.334	2.665	15773882	14063486	7.301	6.949
28) SA Decachlor...	8.760	7.790	21848788	28265744	14.331	14.756

Target Compounds



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL030223\
 Data File : PL081226.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 03 Mar 2023 04:06
 Operator : AR\AJ
 Sample : 01749-16
 Misc :
 ALS Vial : 47 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 CF-18

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Mar 03 04:50:45 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL030223.M
 Quant Title : GC Extractables
 QLast Update : Fri Mar 03 01:48:45 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm






#1 Tetrachloro-m-xylene


R.T.: 3.334 min
 Delta R.T.: -0.001 min
 Response: 15773882
 Conc: 7.30 ng/ml

Instrument :
 ECD_L
 ClientSampleId :
 CF-18




#1 Tetrachloro-m-xylene

R.T.: 2.665 min
 Delta R.T.: -0.001 min
 Response: 14063486
 Conc: 6.95 ng/ml



#28 Decachlorobiphenyl

R.T.: 8.760 min
 Delta R.T.: -0.004 min
 Response: 21848788
 Conc: 14.33 ng/ml



#28 Decachlorobiphenyl

R.T.: 7.790 min
 Delta R.T.: -0.004 min
 Response: 28265744
 Conc: 14.76 ng/ml