

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL030320\
 Data File : PL056968.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 03 Mar 2020 21:05
 Operator : AJ\MA
 Sample : PSTDICC100
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 PSTDICC100

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Mar 04 03:01:42 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL030320.M
 Quant Title : GC Extractables
 QLast Update : Wed Mar 04 02:56:27 2020
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.461	3.958	878.6E6	217.5E6	103.302	104.184
28)	SA Decachlor...	8.215	8.984	952.2E6	208.4E6	101.174	101.446
Target Compounds							
2)	A alpha-BHC	3.898	4.345	1558.5E6	343.4E6	104.728	108.698
3)	MA gamma-BHC...	4.182	4.628	1424.9E6	328.3E6	104.121	107.201
4)	MA Heptachlor	4.483	5.130	1409.2E6	318.7E6	103.502	105.063
5)	MB Aldrin	4.731	5.432	1371.2E6	314.9E6	104.889	106.681
6)	B beta-BHC	4.434	4.800	536.1E6	138.9E6	102.844	102.469
7)	B delta-BHC	4.638	5.014	1395.5E6	280.9E6	105.111	110.625
8)	B Heptachlo...	5.174	5.817	1252.3E6	285.3E6	103.405	105.845
9)	A Endosulfan I	5.513	6.172	1177.9E6	284.8E6	103.265	105.498
10)	B gamma-Chl...	5.401	6.051	1287.1E6	308.6E6	103.710	105.324
11)	B alpha-Chl...	5.460	6.122	1262.5E6	308.0E6	102.857	104.984
12)	B 4,4'-DDE	5.624	6.281	1244.9E6	300.7E6	105.202	107.593
13)	MA Dieldrin	5.756	6.427	1282.8E6	309.3E6	104.666	106.835
14)	MA Endrin	6.014	6.645	1145.9E6	267.6E6	104.734	106.348
15)	B Endosulfa...	6.288	6.854	1100.6E6	272.6E6	103.180	103.118
16)	A 4,4'-DDD	6.140	6.771	1028.1E6	253.7E6	105.064	106.392
17)	MA 4,4'-DDT	6.380	7.068	1072.9E6	259.7E6	107.492	106.175
18)	B Endrin al...	6.456	6.978	914.3E6	232.7E6	103.478	103.295
19)	B Endosulfa...	6.669	7.201	1034.7E6	256.9E6	103.651	104.581
20)	A Methoxychlor	6.926	7.527	520.7E6	127.4E6	104.419	102.573
21)	B Endrin ke...	7.160	7.669	1174.0E6	282.3E6	104.013	104.959
22)	Mirex	7.348	8.120	795.4E6	188.1E6	100.323	101.088

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL030320\
Data File : PL056968.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 03 Mar 2020 21:05
Operator : AJ\MA
Sample : PSTDICC100
Misc :
ALS Vial : 5 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampled :
PSTDICC100

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 04 03:01:42 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL030320.M
Quant Title : GC Extractables
QLast Update : Wed Mar 04 02:56:27 2020
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

