

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL030419\
 Data File : PL045445.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 05 Mar 2019 02:59
 Operator : AJ\SJ
 Sample : K1688-14
 Misc :
 ALS Vial : 67 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 JC-03-030119-G

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Mar 05 04:29:58 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL022719.M
 Quant Title : GC Extractables
 QLast Update : Wed Feb 27 17:00:03 2019
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.364	3.982	140.8E6	44215799	20.471	20.647
28) SA Decachlor...	8.048	9.041	134.7E6	38603330	16.092	16.895
Target Compounds						
7) B delta-BHC	0.000	5.017f	0	1119044	N.D.	0.520 #



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL030419\
Data File : PL045445.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 05 Mar 2019 02:59
Operator : AJ\SJ
Sample : K1688-14
Misc :
ALS Vial : 67 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampleId :
JC-03-030119-G

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 05 04:29:58 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL022719.M
Quant Title : GC Extractables
QLast Update : Wed Feb 27 17:00:03 2019
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x0.25µm






#1 Tetrachloro-m-xylene


R.T.: 3.364 min
Delta R.T.: 0.000 min
Response: 140823965
Conc: 20.47 ng/ml

Instrument :
ECD_L
ClientSampleId :
JC-03-030119-G




#1 Tetrachloro-m-xylene

R.T.: 3.982 min
Delta R.T.: 0.000 min
Response: 44215799
Conc: 20.65 ng/ml




#7 delta-BHC

R.T.: 0.000 min
Exp R.T. : 4.520 min
Response: 0
Conc: N.D.



#7 delta-BHC


R.T.: 5.017 min
Delta R.T.: -0.027 min
Response: 1119044
Conc: 0.52 ng/ml



#28 Decachlorobiphenyl

R.T.: 8.048 min
Delta R.T.: 0.000 min
Response: 134651045
Conc: 16.09 ng/ml

Instrument :
ECD_L
ClientSampleId :
JC-03-030119-G



#28 Decachlorobiphenyl

R.T.: 9.041 min
Delta R.T.: 0.002 min
Response: 38603330
Conc: 16.89 ng/ml