

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL030419\
 Data File : PL045400.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 04 Mar 2019 16:10
 Operator : AJ\SJ
 Sample : K1712-19MSD
 Misc :
 ALS Vial : 28 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 Z-2-4MSD

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Mar 04 23:54:40 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL022719.M
 Quant Title : GC Extractables
 QLast Update : Wed Feb 27 17:00:03 2019
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.364	3.983	157.7E6	50236861	22.920	23.458
28)	SA Decachlor...	8.049	9.043	129.4E6	37231144	15.469	16.294
Target Compounds							
2)	A alpha-BHC	3.794	4.373	706.0E6	195.8E6	63.561	64.433
3)	MA gamma-BHC...	4.072	4.658	646.0E6	182.6E6	63.915	64.638
4)	MA Heptachlor	4.359	5.165	663.1E6	187.8E6	62.840	63.119
5)	MB Aldrin	4.600	5.470	613.2E6	172.1E6	63.045	63.704
6)	B beta-BHC	4.324	4.830	261.5E6	79608758	61.005	62.686
7)	B delta-BHC	4.521	5.046	618.6E6	144.8E6	64.719	67.303
8)	B Heptachlo...	5.038	5.856	573.9E6	154.0E6	61.490	64.128
9)	A Endosulfan I	5.370	6.212	541.2E6	148.7E6	60.553	59.722
10)	B gamma-Chl...	5.261	6.090	598.1E6	162.3E6	61.116	60.480
11)	B alpha-Chl...	5.318	6.163	586.7E6	161.8E6	60.508	59.983
12)	B 4,4'-DDE	5.485	6.321	541.9E6	143.2E6	61.730	61.323
13)	MA Dieldrin	5.610	6.469	588.1E6	155.4E6	61.529	60.454
14)	MA Endrin	5.864	6.687	537.6E6	134.2E6	59.694	58.999
15)	B Endosulfa...	6.139	6.897	494.3E6	135.7E6	57.020	57.692
16)	A 4,4'-DDD	5.997	6.811	429.6E6	116.8E6	58.672	60.718
17)	MA 4,4'-DDT	6.232	7.109	443.3E6	117.9E6	59.110	59.755
18)	B Endrin al...	6.305	7.021	393.2E6	111.7E6	53.795	54.661
19)	B Endosulfa...	6.517	7.243	449.2E6	129.8E6	55.264	57.686
20)	A Methoxychlor	6.777	7.567	230.8E6	65222532	52.556	53.138
21)	B Endrin ke...	7.001	7.713	504.8E6	136.0E6	54.790	57.183
22)	Mirex	7.178	8.166	375.2E6	106.1E6	50.426	53.292

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL030419\
Data File : PL045400.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 04 Mar 2019 16:10
Operator : AJ\SJ
Sample : K1712-19MSD
Misc :
ALS Vial : 28 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampled :
Z-2-4MSD

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 04 23:54:40 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL022719.M
Quant Title : GC Extractables
QLast Update : Wed Feb 27 17:00:03 2019
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

