

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL030421\
 Data File : PL065058.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 04 Mar 2021 14:40
 Operator : AR\AJ
 Sample : PSTDCCC050
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 PSTDCCC050

Manual Integrations
 APPROVED

Ankita
 3/5/2021 9:00:00 AM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Mar 04 16:06:56 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL020821.M
 Quant Title : GC Extractables
 QLast Update : Tue Feb 09 04:41:29 2021
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.428	4.100	138.5E6	122.0E6	49.215	56.457
28)	SA Decachlor...	8.150	9.261	116.9E6	98270707	51.988	51.738
Target Compounds							
2)	A alpha-BHC	3.861	4.502	240.6E6	181.5E6	51.878	59.933m
3)	MA gamma-BHC...	4.142	4.795	218.3E6	168.4E6	50.486	59.287
4)	MA Heptachlor	4.439	5.307	195.4E6	156.8E6	47.668	52.929
5)	MB Aldrin	4.684	5.617	186.5E6	139.9E6	50.074	57.065
6)	B beta-BHC	4.393	4.973	87435533	75018836	51.156	57.895m
7)	B delta-BHC	4.595	5.196	212.6E6	155.9E6	52.595	57.694
8)	B Heptachlo...	5.125	6.011	174.6E6	131.1E6	50.296	59.424
9)	A Endosulfan I	5.459	6.374	158.5E6	116.9E6	49.410m	55.722
10)	B gamma-Chl...	5.350	6.247	179.9E6	129.4E6	51.293	55.988
11)	B alpha-Chl...	5.408	6.319	178.0E6	128.9E6	51.404	55.987
12)	B 4,4'-DDE	5.572	6.476	158.6E6	113.7E6	50.687m	56.046
13)	MA Dieldrin	5.702	6.635	167.6E6	117.8E6	51.098	55.093
14)	MA Endrin	5.956	6.857	142.3E6	82434811	47.383	45.600
15)	B Endosulfa...	6.229	7.072	146.8E6	98901635	50.969	52.486
16)	A 4,4'-DDD	6.087	6.976	147.9E6	101.2E6	58.395	58.386
17)	MA 4,4'-DDT	6.325	7.275	91897971	62774056	40.753	41.172
18)	B Endrin al...	6.397	7.197	122.7E6	81533683	50.320	50.071
19)	B Endosulfa...	6.611	7.421	142.9E6	95807148	50.395	53.238
20)	A Methoxychlor	6.873	7.737	58615473	43109392	45.268	47.690
21)	B Endrin ke...	7.096	7.900	148.1E6	97162218	51.624	50.825
22)	Mirex	7.278	8.351	102.2E6	68324318	48.380	49.153

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL030421\
Data File : PL065058.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 04 Mar 2021 14:40
Operator : AR\AJ
Sample : PSTDCCC050
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampled :
PSTDCCC050

Manual Integrations
APPROVED

Ankita
3/5/2021 9:00:00 AM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 04 16:06:56 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL020821.M
Quant Title : GC Extractables
QLast Update : Tue Feb 09 04:41:29 2021
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

