

Method Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\
 Method File : LTX030521.M
 Title : GC Extractables
 Last Update : Fri Mar 05 02:44:14 2021
 Response Via : Initial Calibration

Calibration Files

500 =PL065077.D 1000 =PL065075.D 250 =PL065078.D
 750 =PL065076.D 100 =PL065079.D

	Compound	500	1000	250	750	100	Avg		%RSD

1) SA	Tetrachloro-m-xylene	3.061	2.893	2.935	3.013	3.090	2.998	E6	2.77
2)	Toxaphene-1	2.473	2.308	2.391	2.359	2.746	2.456	E4	7.05
3)	Toxaphene-2	2.297	2.202	2.207	2.265	2.314	2.257	E4	2.27
4)	Toxaphene-3	5.878	5.699	6.134	5.855	6.795	6.072	E4	7.13
5)	Toxaphene-4	3.992	4.134	3.649	4.141	3.499	3.883	E4	7.54
6)	Toxaphene-5	6.223	6.175	6.222	6.182	6.568	6.274	E4	2.64
7) SA	Decachlorobiphenyl	2.722	2.572	2.776	2.652	3.081	2.761	E6	7.06

Signal #2 Calibration Files

500 =PL065077.D 1000 =PL065075.D 250 =PL065078.D
 750 =PL065076.D 100 =PL065079.D

	Compound	500	1000	250	750	100	Avg		%RSD

1) SA	Tetrachloro-m-xylene	2.532	2.447	2.423	2.524	2.508	2.487	E6	1.96
2)	Toxaphene-1	1.823	1.823	1.792	1.826	1.964	1.845	E4	3.67
3)	Toxaphene-2	1.895	1.878	1.804	1.859	1.856	1.858	E4	1.84
4)	Toxaphene-3	3.716	3.652	3.564	3.686	3.826	3.689	E4	2.59
5)	Toxaphene-4	3.429	3.431	3.319	3.456	3.259	3.379	E4	2.52
6)	Toxaphene-5	3.699	3.609	3.534	3.696	3.491	3.606	E4	2.60
7) SA	Decachlorobiphenyl	2.166	2.041	2.207	2.117	2.339	2.174	E6	5.11

(#) = Out of Range