

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL030521\
 Data File : PL065073.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 04 Mar 2021 21:51
 Operator : AR\AJ
 Sample : PCHLORICC250
 Misc :
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Mar 05 00:44:41 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL030521.M
 Quant Title : GC Extractables
 QLast Update : Fri Mar 05 00:38:45 2021
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.422	4.098	79844414	61149756	24.230	24.879
28)	SA Decachlor...	8.142	9.256	70738148	55475252	26.516	25.978
Target Compounds							
23)	Chlordane-1	4.281	5.111	32242957	28185074	260.217	265.298
24)	Chlordane-2	4.779	5.588	35651904	29095811	266.782	261.821
25)	Chlordane-3	5.343	6.244	97863820	98450137	246.815	244.471
26)	Chlordane-4	5.400	6.317	83520238	120.3E6	252.025	247.022
27)	Chlordane-5	6.229	7.136	32304381	18489832	246.222	240.380



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL030521\
Data File : PL065073.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 04 Mar 2021 21:51
Operator : AR\AJ
Sample : PCHLORICC250
Misc :
ALS Vial : 13 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampled :

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 05 00:44:41 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL030521.M
Quant Title : GC Extractables
QLast Update : Fri Mar 05 00:38:45 2021
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm






#1 Tetrachloro-m-xylene


R.T.: 3.422 min
Delta R.T.: 0.000 min
Response: 79844414
Conc: 24.23 ng/ml

Instrument :
ECD_L
ClientSampleId :




#1 Tetrachloro-m-xylene

R.T.: 4.098 min
Delta R.T.: 0.000 min
Response: 61149756
Conc: 24.88 ng/ml




#23 Chlordane-1

R.T.: 4.281 min
Delta R.T.: 0.000 min
Response: 32242957
Conc: 260.22 ng/ml



#23 Chlordane-1


R.T.: 5.111 min
Delta R.T.: 0.000 min
Response: 28185074
Conc: 265.30 ng/ml



#24 Chlordane-2


R.T.: 4.779 min
Delta R.T.: 0.000 min
Response: 35651904
Conc: 266.78 ng/ml

Instrument :
ECD_L
ClientSampleId :




#24 Chlordane-2

R.T.: 5.588 min
Delta R.T.: 0.000 min
Response: 29095811
Conc: 261.82 ng/ml




#25 Chlordane-3

R.T.: 5.343 min
Delta R.T.: 0.000 min
Response: 97863820
Conc: 246.82 ng/ml



#25 Chlordane-3


R.T.: 6.244 min
Delta R.T.: 0.000 min
Response: 98450137
Conc: 244.47 ng/ml



#26 Chlordane-4


R.T.: 5.400 min
Delta R.T.: 0.000 min
Response: 83520238
Conc: 252.02 ng/ml

Instrument :
ECD_L
ClientSampleId :




#26 Chlordane-4

R.T.: 6.317 min
Delta R.T.: 0.000 min
Response: 120341482
Conc: 247.02 ng/ml




#27 Chlordane-5

R.T.: 6.229 min
Delta R.T.: 0.000 min
Response: 32304381
Conc: 246.22 ng/ml



#27 Chlordane-5


R.T.: 7.136 min
Delta R.T.: 0.000 min
Response: 18489832
Conc: 240.38 ng/ml



#28 Decachlorobiphenyl

R.T.: 8.142 min
Delta R.T.: 0.000 min
Response: 70738148
Conc: 26.52 ng/ml

Instrument :
ECD_L
ClientSampleId :



#28 Decachlorobiphenyl

R.T.: 9.256 min
Delta R.T.: 0.000 min
Response: 55475252
Conc: 25.98 ng/ml