

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL030621\
 Data File : PL065102.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 05 Mar 2021 14:22
 Operator : AR\AJ
 Sample : PSTDCCC050
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 PSTDCCC050

Manual Integrations
 APPROVED

Ankita
 3/8/2021 11:19:02 AM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Mar 05 22:12:37 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL030521.M
 Quant Title : GC Extractables
 QLast Update : Fri Mar 05 02:26:14 2021
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.421	4.098	161.7E6	130.2E6	54.985m	52.439
28)	SA Decachlor...	8.143	9.256	131.7E6	107.4E6	47.449	48.902
Target Compounds							
2)	A alpha-BHC	3.855	4.501	259.3E6	184.8E6	55.014	54.289
3)	MA gamma-BHC...	4.135	4.793	232.7E6	169.0E6	53.769	53.483
4)	MA Heptachlor	4.432	5.303	230.0E6	173.4E6	52.295	51.286
5)	MB Aldrin	4.677	5.613	213.9E6	151.1E6	52.373	51.154
6)	B beta-BHC	4.386	4.972	93683001	79209365	50.158	51.805
7)	B delta-BHC	4.587	5.193	227.9E6	154.0E6	52.202	50.839
8)	B Heptachlo...	5.117	6.008	195.5E6	128.9E6	51.433	53.241
9)	A Endosulfan I	5.452	6.371	182.2E6	128.5E6	50.938	50.599
10)	B gamma-Chl...	5.343	6.245	199.0E6	141.2E6	50.807	50.349
11)	B alpha-Chl...	5.400	6.316	192.9E6	139.3E6	50.338	50.385
12)	B 4,4'-DDE	5.567	6.474	180.5E6	126.1E6	50.232	50.460
13)	MA Dieldrin	5.694	6.631	186.2E6	129.3E6	50.350	50.472
14)	MA Endrin	5.949	6.853	162.0E6	103.2E6	47.183	48.866
15)	B Endosulfa...	6.223	7.069	156.0E6	103.8E6	49.696	48.928
16)	A 4,4'-DDD	6.080	6.973	147.5E6	101.1E6	49.565	50.131
17)	MA 4,4'-DDT	6.318	7.272	135.0E6	96496554	48.432	49.233
18)	B Endrin al...	6.389	7.194	121.7E6	86433822	47.462m	48.854
19)	B Endosulfa...	6.603	7.416	159.6E6	107.0E6	47.220	49.121m
20)	A Methoxychlor	6.864	7.733	73449824	54044330	46.525m	48.760
21)	B Endrin ke...	7.088	7.897	168.6E6	109.1E6	48.845m	51.056
22)	Mirex	7.271	8.347	129.5E6	84587833	48.308	50.013

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL030621\
Data File : PL065102.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 05 Mar 2021 14:22
Operator : AR\AJ
Sample : PSTDCCC050
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampleId :
PSTDCCC050

Manual Integrations
APPROVED

Ankita
3/8/2021 11:19:02 AM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 05 22:12:37 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL030521.M
Quant Title : GC Extractables
QLast Update : Fri Mar 05 02:26:14 2021
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

