

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL030623\
 Data File : PL081269.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 06 Mar 2023 15:49
 Operator : AR\AJ
 Sample : 01787-03MSD
 Misc :
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 COMP-3MSD

Manual Integrations
 APPROVED

Reviewed By : Abdul Mirza 03/07/2023
 Supervised By : Ankita Jodhani 03/07/2023

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Mar 06 20:42:41 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL030223.M
 Quant Title : GC Extractables
 QLast Update : Fri Mar 03 01:48:45 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.334	2.665	36388241	30844940	16.842	15.241
28) SA Decachlor...	8.759	7.789	22566258	28688025	14.802	14.976
Target Compounds						
2) A alpha-BHC	3.784	3.166	159.4E6	136.4E6	50.116	46.546
3) MA gamma-BHC...	4.113	3.494	149.4E6	136.2E6	50.503	49.515
4) MA Heptachlor	4.693	3.828	134.2E6	126.6E6	48.633	45.888
5) MB Aldrin	5.030	4.106	137.9E6	132.0E6	54.213	50.344
6) B beta-BHC	4.315	3.793	64157738	57242166	46.610	49.320
7) B delta-BHC	4.557	4.021	152.0E6	121.5E6	54.678	47.709
8) B Heptachlo...	5.456	4.608	112.5E6	113.3E6	47.685	46.856
9) A Endosulfan I	5.835	4.977	139.2E6	112.5E6	65.866	49.712
10) B gamma-Chl...	5.710	4.856	117.8E6	124.7E6	51.608	50.953m
11) B alpha-Chl...	5.788	4.921	115.4E6	123.8E6	51.221	50.550
12) B 4,4'-DDE	5.966	5.112	149.6E6	158.5E6	76.452	80.748
13) MA Dieldrin	6.111	5.241	109.5E6	114.6E6	52.303	50.819
14) MA Endrin	6.338	5.516	90641985	98228363	49.872	48.879
15) B Endosulfa...	6.559	5.812	92947106	94954819	53.311	51.502
16) A 4,4'-DDD	6.481	5.667	87032367	86666630	56.537	56.493
17) MA 4,4'-DDT	6.793	5.915	75514039	78563882	47.574	48.971
18) B Endrin al...	6.688	5.992	73320282	78185630	51.225m	51.606
19) B Endosulfa...	6.923	6.215	87584209	87142264	48.780	45.338
20) A Methoxychlor	7.273	6.496	42755073	44128656	49.636	47.548
21) B Endrin ke...	7.404	6.721	95731535	100.8E6	52.774	46.825
22) Mirex	7.869	6.893	70805349	84186317	48.526	43.741

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL030623\
Data File : PL081269.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 06 Mar 2023 15:49
Operator : AR\AJ
Sample : 01787-03MSD
Misc :
ALS Vial : 17 Sample Multiplier: 1

Instrument :

ECD_L

ClientSampleId :

COMP-3MSD

Manual Integrations

APPROVED

Reviewed By :Abdul Mirza 03/07/2023

Supervised By :Ankita Jodhani 03/07/2023

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 06 20:42:41 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL030223.M
Quant Title : GC Extractables
QLast Update : Fri Mar 03 01:48:45 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

