

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL030623\
 Data File : PL081260.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 06 Mar 2023 10:44
 Operator : AR\AJ
 Sample : 01755-01
 Misc :
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 BI-WEEKLY-INFLUENT

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Mar 06 20:40:20 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL030223.M
 Quant Title : GC Extractables
 QLast Update : Fri Mar 03 01:48:45 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.336	2.668	41638888	38445498	19.272	18.996
28) SA Decachlor...	8.760	7.790	34019251	30468777	22.314	15.906 #

Target Compounds



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL030623\
Data File : PL081260.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 06 Mar 2023 10:44
Operator : AR\AJ
Sample : 01755-01
Misc :
ALS Vial : 10 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampleId :
BI-WEEKLY-INFLUENT

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 06 20:40:20 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL030223.M
Quant Title : GC Extractables
QLast Update : Fri Mar 03 01:48:45 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm






#1 Tetrachloro-m-xylene


R.T.: 3.336 min
Delta R.T.: 0.001 min
Response: 41638888
Conc: 19.27 ng/ml

Instrument :
ECD_L
ClientSampleId :
BI-WEEKLY-INFLUENT




#1 Tetrachloro-m-xylene

R.T.: 2.668 min
Delta R.T.: 0.002 min
Response: 38445498
Conc: 19.00 ng/ml



#28 Decachlorobiphenyl

R.T.: 8.760 min
Delta R.T.: -0.004 min
Response: 34019251
Conc: 22.31 ng/ml



#28 Decachlorobiphenyl

R.T.: 7.790 min
Delta R.T.: -0.004 min
Response: 30468777
Conc: 15.91 ng/ml