

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL030623\
 Data File : PL081266.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 06 Mar 2023 15:07
 Operator : AR\AJ
 Sample : 01787-02
 Misc :
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 COMP-2

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Mar 06 20:41:59 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL030223.M
 Quant Title : GC Extractables
 QLast Update : Fri Mar 03 01:48:45 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.335	2.665	40173394	36544180	18.594	18.057
28) SA Decachlor...	8.761	7.790	23161835	26992681	15.193	14.091

Target Compounds



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL030623\
Data File : PL081266.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 06 Mar 2023 15:07
Operator : AR\AJ
Sample : 01787-02
Misc :
ALS Vial : 14 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampleId :
COMP-2

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 06 20:41:59 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL030223.M
Quant Title : GC Extractables
QLast Update : Fri Mar 03 01:48:45 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm






#1 Tetrachloro-m-xylene


R.T.: 3.335 min
Delta R.T.: 0.000 min
Response: 40173394
Conc: 18.59 ng/ml

Instrument :
ECD_L
ClientSampleId :
COMP-2




#1 Tetrachloro-m-xylene

R.T.: 2.665 min
Delta R.T.: -0.001 min
Response: 36544180
Conc: 18.06 ng/ml



#28 Decachlorobiphenyl

R.T.: 8.761 min
Delta R.T.: -0.003 min
Response: 23161835
Conc: 15.19 ng/ml



#28 Decachlorobiphenyl

R.T.: 7.790 min
Delta R.T.: -0.004 min
Response: 26992681
Conc: 14.09 ng/ml