

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL030719\
 Data File : PL045602.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 07 Mar 2019 23:12
 Operator : AJ\SJ
 Sample : PSTDCCC050
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 PSTDCCC050

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Mar 08 03:47:54 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL022719.M
 Quant Title : GC Extractables
 QLast Update : Wed Feb 27 17:00:03 2019
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.363	3.982	330.4E6	103.7E6	48.035	48.433
28)	SA Decachlor...	8.044	9.038	391.9E6	110.5E6	46.840	48.359
Target Compounds							
2)	A alpha-BHC	3.793	4.371	544.9E6	150.3E6	49.061	49.478
3)	MA gamma-BHC...	4.070	4.656	498.7E6	139.8E6	49.342	49.481
4)	MA Heptachlor	4.358	5.163	505.0E6	142.3E6	47.860	47.842
5)	MB Aldrin	4.599	5.467	482.0E6	133.3E6	49.557	49.335
6)	B beta-BHC	4.323	4.828	204.8E6	63000477	47.777	49.608
7)	B delta-BHC	4.519	5.044	480.4E6	126.3E6	50.262	58.711
8)	B Heptachlor...	5.036	5.853	460.2E6	128.9E6	49.307	53.647
9)	A Endosulfan I	5.368	6.209	435.8E6	120.7E6	48.754	48.462
10)	B gamma-Chl...	5.259	6.087	474.4E6	131.2E6	48.469	48.890
11)	B alpha-Chl...	5.316	6.160	466.4E6	130.5E6	48.095	48.383
12)	B 4,4'-DDE	5.483	6.317	436.3E6	115.0E6	49.705	49.224
13)	MA Dieldrin	5.608	6.466	471.5E6	125.1E6	49.336	48.699
14)	MA Endrin	5.861	6.683	395.9E6	98659541	43.961	43.361
15)	B Endosulfa...	6.135	6.893	419.2E6	112.5E6	48.360	47.831
16)	A 4,4'-DDD	5.994	6.807	376.4E6	99414857	51.404	51.695
17)	MA 4,4'-DDT	6.229	7.105	328.8E6	87284669	43.841	44.230
18)	B Endrin al...	6.302	7.017	352.8E6	98517177	48.267	48.193
19)	B Endosulfa...	6.514	7.240	392.9E6	107.3E6	48.343	47.699
20)	A Methoxychlor	6.774	7.564	182.4E6	52261571	41.534	42.578
21)	B Endrin ke...	6.997	7.710	472.8E6	121.6E6	51.323	51.131
22)	Mirex	7.174	8.162	347.3E6	96099224	46.683	48.250

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL030719\
Data File : PL045602.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 07 Mar 2019 23:12
Operator : AJ\SJ
Sample : PSTDCCC050
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampleId :
PSTDCCC050

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 08 03:47:54 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL022719.M
Quant Title : GC Extractables
QLast Update : Wed Feb 27 17:00:03 2019
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

