

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL030722\
 Data File : PL073180.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 07 Mar 2022 15:27
 Operator : AR\AJ
 Sample : N1733-14
 Misc :
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Mar 08 04:50:52 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL022522CLP.M
 Quant Title : GC Extractables
 QLast Update : Tue Mar 01 01:47:07 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml


System Monitoring Compounds							
1)	SA Tetrachlo...	4.645	5.492	5662951	27388341	16.467	17.544
27)	SA Decachlor...	10.426	11.332	16446673	54033710	34.206	33.430
Target Compounds							
2)	A alpha-BHC	0.000	6.027f	0	399537	N.D.	0.207
4)	MA Heptachlor	0.000	7.105	0	281866	N.D.	0.139
5)	MB Aldrin	6.518	0.000	419467	0	0.927	N.D. #
6)	B beta-BHC	0.000	6.668	0	1281908	N.D.	1.310
7)	B delta-BHC	0.000	6.944	0	3928478	N.D.	2.032
8)	B Heptachlo...	0.000	7.966f	0	538928	N.D.	0.317
9)	A Endosulfan I	0.000	8.363f	0	30702	N.D.	<MDL
15)	B Endosulfa...	8.323f	9.009f	1229532	95313	2.624	<MDL #
21)	B Endrin ke...	0.000	9.953	0	59668	N.D.	<MDL
23)	Toxaphene-2	0.000	9.009	0	95313	N.D.	4.554


(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL030722\
Data File : PL073180.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 07 Mar 2022 15:27
Operator : AR\AJ
Sample : N1733-14
Misc :
ALS Vial : 17 Sample Multiplier: 1

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 08 04:50:52 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL022522CLP.M
Quant Title : GC Extractables
QLast Update : Tue Mar 01 01:47:07 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm






#1 Tetrachloro-m-xylene


R.T.: 4.645 min
Delta R.T.: 0.000 min
Response: 5662951
Conc: 16.47 ng/ml

Instrument :
ECD_L
ClientSampleId :




#1 Tetrachloro-m-xylene

R.T.: 5.492 min
Delta R.T.: 0.003 min
Response: 27388341
Conc: 17.54 ng/ml




#2 alpha-BHC

R.T.: 0.000 min
Exp R.T. : 5.355 min
Response: 0
Conc: N.D.



#2 alpha-BHC


R.T.: 6.027 min
Delta R.T.: -0.025 min
Response: 399537
Conc: 0.21 ng/ml



#4 Heptachlor


R.T.: 0.000 min
Exp R.T.: 6.192 min
Response: 0
Conc: N.D.

Instrument :
ECD_L
ClientSampleId :




#4 Heptachlor

R.T.: 7.105 min
Delta R.T.: 0.005 min
Response: 281866
Conc: 0.14 ng/ml




#5 Aldrin

R.T.: 6.518 min
Delta R.T.: 0.000 min
Response: 419467
Conc: 0.93 ng/ml



#5 Aldrin


R.T.: 0.000 min
Exp R.T.: 7.466 min
Response: 0
Conc: N.D.



#6 beta-BHC


R.T.: 0.000 min
Exp R.T.: 6.158 min
Response: 0
Conc: N.D.

Instrument :
ECD_L
ClientSampleId :




#6 beta-BHC

R.T.: 6.668 min
Delta R.T.: -0.007 min
Response: 1281908
Conc: 1.31 ng/ml




#7 delta-BHC

R.T.: 0.000 min
Exp R.T.: 6.420 min
Response: 0
Conc: N.D.



#7 delta-BHC


R.T.: 6.944 min
Delta R.T.: 0.000 min
Response: 3928478
Conc: 2.03 ng/ml



#8 Heptachlor epoxide


R.T.: 0.000 min
Exp R.T.: 7.094 min
Response: 0
Conc: N.D.

Instrument :
ECD_L
ClientSampleId :




#8 Heptachlor epoxide

R.T.: 7.966 min
Delta R.T.: 0.044 min
Response: 538928
Conc: 0.32 ng/ml




#15 Endosulfan II

R.T.: 8.323 min
Delta R.T.: -0.063 min
Response: 1229532
Conc: 2.62 ng/ml



#15 Endosulfan II


R.T.: 9.009 min
Delta R.T.: -0.054 min
Response: 95313
Conc: N.D.



#23 Toxaphene-2


R.T.: 0.000 min
Exp R.T.: 7.925 min
Response: 0
Conc: N.D.

Instrument :
ECD_L
ClientSampleId :




#23 Toxaphene-2

R.T.: 9.009 min
Delta R.T.: -0.014 min
Response: 95313
Conc: 4.55 ng/ml



#27 Decachlorobiphenyl

R.T.: 10.426 min
Delta R.T.: -0.001 min
Response: 16446673
Conc: 34.21 ng/ml



#27 Decachlorobiphenyl

R.T.: 11.332 min
Delta R.T.: 0.008 min
Response: 54033710
Conc: 33.43 ng/ml