

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL031120\
 Data File : PL057249.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 12 Mar 2020 01:38
 Operator : AJ\MA
 Sample : PB127460BS
 Misc :
 ALS Vial : 56 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 PB127460BS

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Mar 12 06:50:10 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL030320.M
 Quant Title : GC Extractables
 QLast Update : Wed Mar 11 17:44:01 2020
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.460	3.958	165.4E6	40630159	19.405	19.044
28)	SA Decachlor...	8.207	8.982	161.7E6	38042712	16.734	17.905
Target Compounds							
2)	A alpha-BHC	3.897	4.345	765.0E6	155.4E6	51.619	49.191
3)	MA gamma-BHC...	4.180	4.627	694.1E6	149.3E6	50.513	48.031m
4)	MA Heptachlor	4.481	5.131	712.9E6	150.8E6	52.482	48.461
5)	MB Aldrin	4.729	5.433	673.2E6	150.4E6	52.131	50.299
6)	B beta-BHC	4.432	4.801	289.2E6	71452696	54.588	50.698
7)	B delta-BHC	4.636	5.015	667.4E6	125.3E6	51.155	51.472
8)	B Heptachlo...	5.171	5.817	622.1E6	136.9E6	51.414	49.397
9)	A Endosulfan I	5.510	6.172	588.8E6	134.0E6	51.944	47.976
10)	B gamma-Chl...	5.398	6.051	636.5E6	145.8E6	51.244	48.709
11)	B alpha-Chl...	5.456	6.123	618.0E6	142.9E6	50.301	47.398
12)	B 4,4'-DDE	5.620	6.281	605.9E6	135.1E6	51.704	47.592
13)	MA Dieldrin	5.752	6.428	623.1E6	138.8E6	50.960	47.161
14)	MA Endrin	6.010	6.644	514.7E6	108.4E6	47.309	42.046
15)	B Endosulfa...	6.283	6.854	509.1E6	115.3E6	47.140	41.077
16)	A 4,4'-DDD	6.136	6.770	494.9E6	112.3E6	51.119	46.113
17)	MA 4,4'-DDT	6.376	7.067	422.0E6	94261479	43.130	38.318
18)	B Endrin al...	6.451	6.978	430.6E6	102.3E6	48.127	44.168
19)	B Endosulfa...	6.664	7.201	483.7E6	114.3E6	48.437	45.578
20)	A Methoxychlor	6.921	7.527	210.1E6	51495275	42.140	41.219
21)	B Endrin ke...	7.154	7.668	534.5E6	123.1E6	47.334	44.919
22)	Mirex	7.342	8.118	366.3E6	87057480	45.200	45.410

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL031120\
Data File : PL057249.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 12 Mar 2020 01:38
Operator : AJ\MA
Sample : PB127460BS
Misc :
ALS Vial : 56 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampled :
PB127460BS

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 12 06:50:10 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL030320.M
Quant Title : GC Extractables
QLast Update : Wed Mar 11 17:44:01 2020
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

